

User: Huhn, William
 Westchester Dept. of Labs & Research
 Date: 10/17/2001 Time: 15:13:10

Sample: AD25051
 Analysis: \$F_524 Volatile Organic Compounds-Field Blank
 Units: ug
 Started: 10/15/01 00:07 Ended: 10/15/01 00:07 Analyst: BH
 Result source: manual entry
 Page: 1

	Component Name	Vol	Result	Qualifier	MDL
1	Surr_1,2-DICHLOROETHANE-		5.3		.5
2	Surr_4-BROMOFLUOROBENZE		4.1		.5
3	Dichlorodifluoromethane		< MDL		.5
4	Chloromethane		< MDL		.5
5	Vinyl chloride		< MDL		.5
6	Bromomethane		< MDL		.5
7	Chloroethane		< MDL		.5
8	Trichlorofluoromethane		< MDL		.5
9	1,1-Dichloroethene		< MDL		.5
10	Methylene Chloride		< MDL		.5
11	Methyl tert-butyl ether (MTBE)		< MDL		1.0
12	trans-1,2-dichloroethene		< MDL		.5
13	1,1-dichloroethane		< MDL		.5
14	2-butanone (MEK)		< MDL		2.0
15	2,2-dichloropropane		< MDL		.5
16	cis-1,2-dichloroethene		< MDL		.5
17	Chloroform		< MDL		.5
18	Bromochloromethane		< MDL		.5
19	1,2-dichloroethane		< MDL		.5
20	Surr_TOLUENE-D8		5.0		.5
21	1,1,1- trichloroethane		< MDL		.5
22	1,1-dichloropropene		< MDL		.5
23	Carbon tetrachloride		< MDL		.5
24	Benzene		< MDL		.5
25	Trichloroethene		< MDL		.5
26	1,2-dichloropropane		< MDL		.5
27	Dibromomethane		< MDL		.5
28	Bromodichloromethane		< MDL		.5
29	Methy iso-butyl ketone (MIBK)		< MDL		1.0
30	cis-1,3-dichloropropene		< MDL		.5
31	Toluene		< MDL		.5
32	trans-1,3-dichloropropene		< MDL		.5
33	1,1,2-trichloroethane		< MDL		.5
34	Tetrachloroethene		< MDL		.5
35	1,3-dichloropropane		< MDL		.5
36	Dibromochloromethane		< MDL		2.0
37	Chlorobenzene		< MDL		.5
38	1,1,1,2-tetrachloroethane		< MDL		.5
39	Ethylbenzene		< MDL		.5
40	P & M-xylene		< MDL		.5
41	O-xylene		< MDL		.5
42	Styrene		< MDL		.5
43	1,1,2,2-tetrachloroethane		< MDL		.5
44	Bromoform		< MDL		2.0
45	Isopropylbenzene		< MDL		.5
46	1,2,3-trichloropropane		< MDL		.5
47	N-propylbenzene		< MDL		.5
48	Bromobenzene		< MDL		.5

User: Huhn, William
Westchester Dept. of Labs & Research
Date: 10/17/2001 Time: 15:13:10

Sample: AD25051
Analysis: \$F_524 Volatile Organic Compounds-Field Blank
Units: ug
Started: 10/15/01 00:07 Ended: 10/15/01 00:07 Analyst: BH
Result source: manual entry
Page: 2

	Component Name	Viol	Result	Qualifier	MDL
49	1,3,5-trimethylbenzene		< MDL		.5
50	2-chlorotoluene		< MDL		.5
51	4-chlorotoluene		< MDL		.5
52	TERT-butylbenzene		< MDL		.5
53	1,2,4-trimethylbenzene		< MDL		.5
54	SEC-butylbenzene		< MDL		.5
55	P-isopropyltoluene		< MDL		.5
56	1,3-dichlorobenzene		< MDL		.5
57	1,4-dichlorobenzene		< MDL		.5
58	N-butylbenzene		< MDL		.5
59	1,2-dichlorobenzene		< MDL		.5
60	1,2,4-trichlorobenzene		< MDL		.5
61	Hexachlobutadiene		< MDL		.5
62	Naphthalene		< MDL	0.63	.5
63	1,2,3-trichlorobenzene		< MDL	0.11	.5

Quantitation Report

(Not Reviewed)

Data File : C:\HPCHEM\1\DATA\01OCT15\25051F.D
 Acq On : 15 Oct 2001 7:21 pm
 Sample : nyc fb
 Misc :
 MS Integration Params: LSCINT.P
 Quant Time: Oct 15 20:00 2001

Vial: 11
 Operator:
 Inst : GC/MS Ins
 Multiplr: 1.00

Quant Results File: HP092601.RES

Quant Method : C:\HPCHEM\1\METHODS\HP092601.M (RTE Integrator)
 Title : VOCs BY EPA METHOD 524
 Last Update : Wed Sep 26 08:36:04 2001
 Response via : Initial Calibration
 DataAcq Meth : HP092601

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-DIFLUOROBENZENE	14.45	114	996141	5.00	ug/L	-0.10
22) CHLOROBENZENE-D5	21.05	117	940934	5.00	ug/L	-0.10
50) 1,4-DICHLOROBENZENE-D4	26.21	152	427673	5.00	ug/L	-0.10
System Monitoring Compounds						
2) 1,2-DICHLOROETHANE-D4	13.67	65	236426	5.29	ug/L	-0.10
Spiked Amount 5.000			Recovery	=	105.80%	
3) 4-BROMOFLUOROBENZENE	23.62	174	299778	4.09	ug/L	-0.12
Spiked Amount 5.000			Recovery	=	81.80%	
23) TOLUENE-D8	17.78	98	1243144	4.99	ug/L	-0.10
Spiked Amount 5.000			Recovery	=	99.80%	
Target Compounds						
10) ACETONE	7.69	43	80388	21.52	ug/L	89
69) 1,2,4-TRICHLOROBENZENE	30.71	180	11769	0.25	ug/L	98
70) HEXACHLOROBUTADIENE	30.95	225	7969	0.25	ug/L	94
71) NAPHTHALENE	31.37	128	27599	0.53	ug/L	99
72) 1,2,3-TRICHLOROBENZENE	31.93	180	24201	0.71	ug/L	94

do {

(#) = qualifier out of range (m) = manual integration
 25051F.D HP092601.M Tue Nov 13 10:53:04 2001

Quantitation Report

Data File : C:\HPCHEM\1\DATA\01OCT15\25051F.D
Acq On : 15 Oct 2001 7:21 pm
Sample : nyc fb
Misc :
MS Integration Params: LSCINT.P
Quant Time: Oct 15 20:00 2001

Vial: 11
Operator:
Inst : GC/MS Ins
Multiplr: 1.00

Quant Results File: HP092601.RES

Method : M:\INSTRU~1\206\METHODS\HP092601.M (RTE Integrator)
Title : VOCs BY EPA METHOD 524
Last Update : Wed Sep 26 08:36:04 2001
Response via : Initial Calibration

