

DFTPP

Data File : C:\MSDCHEM\1\DATA\022802\DFTPC.D

Acq On : 1 Mar 2002 12:23 pm

Sample :

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 2290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	57.2	302720	PASS
68	69	0.00	2	1.4	3558	PASS
69	198	0.00	100	47.4	250944	PASS
70	69	0.00	2	0.3	819	PASS
127	198	40	60	49.3	260992	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	529664	PASS
199	198	5	9	6.3	33560	PASS
275	198	10	30	20.9	110776	PASS
365	198	1	100	2.6	13817	PASS
441	443	0.01	100	71.4	52672	PASS
442	198	40	110	72.0	381120	PASS
443	442	17	23	19.4	73816	PASS

DFTPC.D 5080202.M

Mon Mar 04 05:59:05 2002

File : C:\MSDCHEM\1\DATA\022802\DFTPC.D
 Operator : McLean
 Acquired : 1 Mar 2002 12:23 pm using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name:
 Misc Info :
 Vial Number: 1

