

Method 508Date Extracted 10/11/01

Page _____

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH = 7	Count		
	Lab 11h 500ml		-	1ml		-	2ml					
	ref chloroform											
	MDL chlor.											
	PRC 1254											
24334	Spike pest				1ml							
24334	Spike Dup pest				↓							
24395	Spike Tox				↓							
24395	Spike Dup Tox											
	24335									18		
	24396									19		
	24397									20		
	24398									1		
	24399									2		
	24400									3		
	24401									4		
	24429									5		
	24430	↓		↓			↓		↓	5		
	24334									5		
	24395	↓		↓			↓		↓	8		
	24333	↓		↓								
Remarks												
24395 Sp Dup Tox - lab accident - lost most of ext.												
MDL chloroform - cancelled - not required												
drip rate must be between 5-15 ml/min for liq-liq extractors												
Extraction type - Sup fume Solv. lot# NaCl lot# Na2SO4 lot#												
STANDARD	INT STD LOG#											
CALIBRATION												
REF/SPIKE	POX 1678a chlor.						Pest 2062a					
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	2064C											
LPC												
ENDRIN/DDT												

Analyst _____
revision 4/12/01

Verified by _____

Reviewed by _____
Date 11/26/01