

Method 508 MOD.

Date Extracted 2-4-02

Page 65

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
L.B.		500 <i>inh</i>		<i>inh</i>			1000	1000	7			
Ref.					0.5ml			1000				
2457								98				
2458								9A				
2459								88				
2460	F.B.							8A				
2561								7A				
2562								7B				
2563								6B				
2564								6A				
2565								4B				
2566								4A				
Remarks												
Reviewed by <i>[Signature]</i> 8/13/02												
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate - Extraction type - <i>SEP. Fur.</i> Solv. lot# <i>V 41470</i> NaCl lot# <i>V 11599</i> Na2SO4 lot# <i>V 39400</i>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF/SPIKE	<i>95759</i>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<i>020102</i>											
LPC												
ENDRIN/DDT												

Analyst A.D.
revision 4/12/01

Verified by _____

Reviewed by *[Signature]*
Date 3/7/02

NYC DEP Discount

AE 02458 low surrogate 64%

DFTPP

• Data File : C:\MSDCHEM\1\DATA\022702\DFTPP.D
Acq On : 27 Feb 2002 6:03 am
Sample : DFTPP
Misc :
MS Integration Params: BENZO.P

Vial: 1
Operator: McLean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)
Title : Quantitation For Method.508gcscan

Spectrum Information: Scan 2291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	52.8	271872	PASS
68	69	0.00	2	1.8	4054	PASS
69	198	0.00	100	43.3	223168	PASS
70	69	0.00	2	0.6	1298	PASS
127	198	40	60	47.2	243328	PASS
197	198	0.00	2	0.5	2396	PASS
198	198	100	100	100.0	515136	PASS
199	198	5	9	6.7	34312	PASS
275	198	10	30	22.5	116000	PASS
365	198	1	100	3.0	15283	PASS
441	443	0.01	100	78.3	61016	PASS
442	198	40	110	81.7	420736	PASS
443	442	17	23	18.5	77920	PASS

DFTPP.D 5080202.M Fri Mar 01 08:45:16 2002