

User: Vinci, David L.
 Westchester Dept. of Labs & Research
 Date: 03/07/2002 Time: 16:39:14

Sample: AE02562
 Analysis: \$GCMS GCMS Scan For Unknowns
 Units: ug
 Started: 3/7/02 16:38 Ended: 3/7/02 16:38 Analyst: DLV
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Other unknown compounds				
2	Surr_2,4-DDD		71 ✓		5.0

Comments for Sample AE02562 - Analysis \$GCMS

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 03-07-2002 Time: 16:39:19

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Quantitation Report (QT Reviewed)

Data File : C:\MSDCHEM\1\DATA\022702\02562.D

Acq On : 27 Feb 2002 4:28 pm

Sample : GC SCAN 2/4

Misc :

MS Integration Params: BENZO.P

Quant Time: Mar 01 08:55:14 2002

Vial: 10

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 5080202.RES

Quant Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Last Update : Mon Feb 25 19:17:14 2002

Response via : Initial Calibration

DataAcq Meth : 508SCAN

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	150	1699045	20.00	ug/L	0.00
10) Naphthalene-d8	13.05	136	4418864	20.00	ug/L	0.00
17) Acenaphthene-d10	18.17	164	2439860	20.00	ug/L	0.00
29) Phenanthrene-d10	22.53	188	3701486	20.00	ug/L	0.00
38) Chrysene-d12	30.35	240	3551343	20.00	ug/L	0.00
44) Perylene-d12	34.22	264	2514915	20.00	ug/L	0.00

System Monitoring Compounds

37) 2,4-ddd surr	27.47	235	207313	3.53	ug/L	0.00
Spiked Amount	5.000		Recovery	=	70.60%	

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) N-nitroso-dimethyl amine	0.00	74	0	N.D.	d	
3) bis(2-chloroethyl) ether	0.00	93	0	N.D.		
4) 1,3 - Dichlorobenze	0.00	146	0	N.D.		
5) 1,4 - Dichlorobenzene	0.00	146	0	N.D.		
6) 1,2 - Dichlorobenzene	0.00	146	0	N.D.		
7) 2,2' - oxybis (1-Chloropr	0.00	45	0	N.D.		
8) N-Nitroso-di-n-propylamine	0.00	70	0	N.D.		
9) Hexachloroethane	0.00	117	0	N.D.		
11) Nitrobenzene	0.00	77	0	N.D.		
12) Isophorone	0.00	82	0	N.D.		
13) bis(2-Chloroethoxy) methan	0.00	93	0	N.D.		
14) 1,2,4-Trichlorobenzene	0.00	180	0	N.D.		
15) Naphthalene	0.00	128	0	N.D.		
16) Hexachlorobutadiene	0.00	225	0	N.D.		
18) Hexachlorocyclopentadiene	0.00	237	0	N.D.		
19) 2-chloronaphthalene	0.00	162	0	N.D.		
20) Dimethylphthalate	0.00	163	0	N.D.		
21) 2,6-Dinitrotoluene	0.00	165	0	N.D.	d	
22) Acenaphthylene	0.00	152	0	N.D.		
23) Acenaphthene	0.00	153	0	N.D.		
24) 2,4-Dinitrotoluene	0.00	165	0	N.D.		
25) Diethylphthalate	0.00	149	0	N.D.		
26) 4-Chlorophenyl-phenyl ethe	0.00	204	0	N.D.		
27) Fluorene	0.00	166	0	N.D.		
28) Azobenzen/ 1,2-Diphenylhyd	0.00	77	0	N.D.		
30) N-Nitrosodiphenylamine	0.00	169	0	N.D.		
31) 4-Bromophenyl-phenyl ether	0.00	248	0	N.D.		
32) Hexachlorobenzene	0.00	284	0	N.D.		
33) Phenanthrene	0.00	178	0	N.D.		
34) Anthracene	0.00	178	0	N.D.		
35) Di-n-butylphthalate	24.68	149	4387	0.02	ug/L	79
36) Fluoranthene	0.00	202	0	N.D.		
39) Pyrene	0.00	202	0	N.D.		
40) Butylbenzylphthalate	0.00	149	0	N.D.		
41) Benzo (a) anthracene	30.35	228	8736	0.04	ug/L	42
42) Bis (2-ethylhexyl) phthala	30.94	149	4254	0.03	ug/L	88
43) Chrysene	30.35	228	8736	0.04	ug/L	40
45) Di-n-Octylphthalate	0.00	149	0	N.D.		
46) Benzo (b) fluoranthene	0.00	252	0	N.D.		

(#)= qualifier out of range (m) = manual integration

02562.D 5080202.M Fri Mar 01 08:55:46 2002

Data File : C:\MSDCHEM\1\DATA\022702\02562.D

Acq On : 27 Feb 2002 4:28 pm

Sample : GC SCAN 2/4

Misc :

MS Integration Params: BENZO.P

Quant Time: Mar 1 8:55 2002

Vial: 10

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 5080202.RES

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Last Update : Mon Feb 25 19:17:14 2002

Response via : Initial Calibration

