

Method 508 mod.

Date Extracted 2-4-02

Page 68

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1ml			1ml	8A	7			
REF					0.5ml			8B				
25 67								9A				
25 68								10A				
25 75								6B				
25 76								9B				
25 77								7B				
25 78								10B				
25 79								7A				
25 80								8B				
25 81 *								8A				
25 82								6A				

Remarks

* Sample # 2581 leaked out. Will be re-extracted on 2/5/02.
(entire batch to be re-expt)

drip rate must be between 5-15 ml/min for liq-liq extractors

drip rate -

Extraction type 3 lb fun Solv. lot# V41470 NaCl lot# V39618 Na2SO4 lot# V39400

STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	91759											
MDL/MDRF												
Q. CHECK												
INT. STD.	022202											
SURROGATE	GCMS	020102										
LPC												
ENDRIN/DDT												

Reviewed by,

C/B 8/3/02

Analyst CH
revision 4/12/01

Verified by

my

Reviewed by
Date

4/2/02
DN