

DFTPP

Data File : C:\MSDCHEM\1\DATA\022702\DFTPPB.D
Acq On : 27 Feb 2002 8:34 pm
Sample :
Misc :
MS Integration Params: BENZO.P

Vial: 1
Operator: McLean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)
Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 2291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	50.2	272768	PASS
68	69	0.00	2	1.6	3559	PASS
69	198	0.00	100	41.9	227584	PASS
70	69	0.00	2	0.3	690	PASS
127	198	40	60	46.0	250304	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	543808	PASS
199	198	5	9	6.1	33096	PASS
275	198	10	30	22.4	122072	PASS
365	198	1	100	3.0	16090	PASS
441	443	0.01	100	74.4	70936	PASS
442	198	40	110	87.9	478080	PASS
443	442	17	23	20.0	95408	PASS

DFTPPB.D 5080202.M Fri Mar 01 09:11:49 2002

File : C:\MSDCHEM\1\DATA\022702\DFTPPB.D
 Operator : McLean
 Acquired : 27 Feb 2002 8:34 pm using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name:
 Misc Info :
 Vial Number: 1

