

User: Fakhouri, Morgan
 Westchester Dept. of Labs & Research
 Date: 03/19/2002 Time: 11:39:15

Sample: AE03075
 Analysis: \$C1GCMS CALI GCMS
 Units: ug
 Started: 3/19/02 11:32 Ended: 3/19/02 11:32 Analyst: MF
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	n-Nitrosodimethylamine		39.65		2.0
2	bis(2-Chloroethyl)ether		37.55		2.0
3	1,3-Dichlorobenzene		37.55		2.0
4	1,4-Dichlorobenzene		37.60		2.0
5	1,2-Dichlorobenzene		37.01		2.0
6	2,2'-oxybis(1-Chloropropane)		39.25		2.0
7	n-Nitrosodi-n-propylamine		40.09		2.0
8	Hexachloroethane		37.88		2.0
9	Nitrobenzene		38.34		2.0
10	Isophorone		38.85		2.0
11	bis(2-Chloroethoxy)methane		37.90		2.0
12	1,2,4-Trichlorobenzene		37.88		2.0
13	Naphthalene		37.44		2.0
14	Hexachlorobutadiene		38.64		2.0
15	2-Chloronaphthalene		38.26		2.0
16	Dimethylphthalate		39.31		2.0
17	2,6-Dinitrotoluene		39.75		2.0
18	Acenaphthylene		38.35		2.0
19	Acenaphthene		38.07		2.0
20	2,4-Dinitrotoluene		40.64		2.0
21	Diethylphthalate		38.54		2.0
22	4-Chlorophenylphenylether		38.38		2.0
23	Fluorene		37.33		2.0
24	Azobenzene/1,2-Diphenylhydraz		40.89		2.0
25	n-Nitrosodiphenylamine		37.95		2.0
26	4-Bromophenylphenylether		37.37		2.0
27	Hexachlorobenzene		37.60		2.0
28	Phenanthrene		37.34		2.0
29	Anthracene		36.81		2.0
30	Di-n-butylphthalate		38.61		2.0
31	Fluoranthene		38.52		2.0
32	Pyrene		36.12		2.0
33	Butylbenzylphthalate		37.23		2.0
34	Benzo(a)anthracene		38.44		2.0
35	bis(2-Ethylhexyl)phthalate		35.60		2.0
36	Chrysene		37.96		2.0
37	Di-n-octylphthalate		37.35		2.0
38	Benzo(b)fluoranthene		39.57		2.0
39	Benzo(k)fluoranthene		39.65		2.0
40	Benzo(a)pyrene		39.09		2.0
41	Indeno(1,2,3-cd)pyrene		40.60		2.0
42	Dibenz(a,h)anthracene		41.14		2.0
43	Benzo(g,h,i)perylene		41.38		2.0

Quantitation Report

(QT Reviewed)

Data File : C:\MSDCHEM\1\DATA\031802\CAL40.D

Acq On : 19 Mar 2002 3:00 am

Sample :

Misc :

MS Integration Params: BENZO.P

Quant Time: Mar 19 21:41:06 2002

Vial: 2

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 5080302.RES

Quant Method : C:\MSDCHEM\1\5973N\5080302.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Last Update : Thu Mar 14 13:02:27 2002

Response via : Initial Calibration

DataAcq Meth : 508SCAN

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.56	150	2699526	20.00	ug/L	0.00
10) Naphthalene-d8	13.04	136	6468500	20.00	ug/L	0.02
17) Acenaphthene-d10	18.15	164	3489303	20.00	ug/L	0.00
29) Phenanthrene-d10	22.51	188	5730838	20.00	ug/L	0.02
38) Chrysene-d12	30.35	240	4783485	20.00	ug/L	0.03
44) Perylene-d12	34.20	264	2739934	20.00	ug/L	0.02

System Monitoring Compounds

37) 2,4-ddd surr	27.44	235	45351	0.66	ug/L	0.00
Spiked Amount	5.000		Recovery	=	13.20%	

Target Compounds

					Qvalue
2) N-nitroso-dimethyl amine	3.57	74	2403604	39.65	100
3) bis(2-chloroethyl) ether	9.03	93	3588007	37.55	94
4) 1,3 - Dichlorobenze	9.40	146	3549376	37.55	99
5) 1,4 - Dichlorobenzene	9.61	146	3544515	37.60	99
6) 1,2 - Dichlorobenzene	9.99	146	3229277	37.01	99
7) 2,2' - oxybis (1-Chloropr	10.45	45	8542684	39.25	99
8) N-Nitroso-di-n-propylamine	10.82	70	2754786	40.09	99
9) Hexachloroethane	10.87	117	1341141	37.88	96
11) Nitrobenzene	11.15	77	4466237	38.34	100
12) Isophorone	11.87	82	8328304	38.85	99
13) bis(2-Chloroethoxy) methan	12.60	93	4889783	37.90	99
14) 1,2,4-Trichlorobenzene	12.93	180	2805528	37.88	98
15) Naphthalene	13.11	128	10065118	37.44	99
16) Hexachlorobutadiene	13.54	225	1600083	38.64	98
18) Hexachlorocyclopentadiene	15.62	237	1359331	44.24	97
19) 2-chloronaphthalene	16.54	162	6093409	38.26	98
20) Dimethylphthalate	17.65	163	7473234	39.31	99
21) 2,6-Dinitrotoluene	17.75	165	1608432	39.75	99
22) Acenaphthylene	17.72	152	8475751	38.35	99
23) Acenaphthene	18.26	153	5719113	38.07	100
24) 2,4-Dinitrotoluene	18.91	165	2355034	40.64	98
25) Diethylphthalate	19.78	149	6307028	38.54	100
26) 4-Chlorophenyl-phenyl ethe	19.92	204	3347013	38.38	97
27) Fluorene	19.79	166	6611417	37.33	98
28) Azobenzen/ 1,2-Diphenylhyd	20.39	77	9868830	40.89	99
30) N-Nitrosodiphenylamine	20.32	169	4674302	37.95	99
31) 4-Bromophenyl-phenyl ether	21.34	248	1890590	37.37	97
32) Hexachlorobenzene	21.38	284	2006137	37.60	99
33) Phenanthrene	22.60	178	9911778	37.34	99
34) Anthracene	22.75	178	9694794	36.81	99
35) Di-n-butylphthalate	24.67	149	11745444	38.61	100
36) Fluoranthene	26.10	202	11616633	38.52	99
39) Pyrene	26.73	202	11946179	36.12	96
40) Butylbenzylphthalate	29.08	149	5270824	37.23	97
41) Benzo (a) anthracene	30.31	228	9650114	38.44	100
42) Bis (2-ethylhexyl) phthala	30.93	149	6664177	35.60	99
43) Chrysene	30.43	228	9054976	37.96	100
45) Di-n-Octylphthalate	32.78	149	10990981	37.35	97
46) Benzo (b) fluoranthene	33.27	252	7640138	39.57	99

(#)=qualifier out of range (m)=manual integration

CAL40.D 5080302.M Tue Mar 19 21:41:14 2002

Quantitation Report

(QT Reviewed)

Data File : C:\MSDCHEM\1\DATA\031802\CAL40.D

Acq On : 19 Mar 2002 3:00 am

Sample :

Misc :

MS Integration Params: BENZO.P

Quant Time: Mar 19 21:41:06 2002

Vial: 2

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 5080302.RES

Quant Method : C:\MSDCHEM\1\5973N\5080302.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Last Update : Thu Mar 14 13:02:27 2002

Response via : Initial Calibration

DataAcq Meth : 508SCAN

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
47) Benzo (k) fluoranthene	33.36	252	7808188	39.65	ug/L	98
48) Benzo (a) pyrene	34.07	252	6447459	39.09	ug/L	99
49) Indeno (1,2,3-cd) pyrene	36.75	276	3852734	40.60	ug/L	94
50) Dibenz (a,h) anthracene	36.86	278	3754752	41.14	ug/L	96
51) Benzo (g,h,i) perylene	37.42	276	3779100	41.38	ug/L	97

(#) = qualifier out of range (m) = manual integration

CAL40.D 5080302.M Tue Mar 19 21:41:14 2002

Page 2

Data File : C:\MSDCHEM\1\DATA\031802\CAL40.D

Acq On : 19 Mar 2002 3:00 am

Sample :

Misc :

MS Integration Params: BENZO.P

Quant Time: Mar 19 21:41 2002

Vial: 2

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 5080302.RES

Method : C:\MSDCHEM\1\5973N\5080302.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Last Update : Thu Mar 14 13:02:27 2002

Response via : Initial Calibration

