

Method 508 MOD.

Date Extracted 5/21/02

Page 136

Ref

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
L.B.		500 mL		1 mL			1 mL	9B	7			
Ref 1					1 mL			3B				
Ref 2					1 mL			2B				
11799								10B				
11800	F.B.							6B				
11901								7B				
11902								9A				
11903								8B				
11904	F.B.							2A				
11905								10B				
11906								7B				
11907								8A				
11908	F.B.							3A				
11909												
11910												
Remarks												
Reviewed by, <u>C/B 8/23/02</u>												
drip rate must be between 5-15 ml/min for liq-liq extractors												
drip rate -												
Extraction type - <u>SEP. Fun.</u> Solv. lot# <u>X07E07</u> NaCl lot# <u>V39614</u> Na2SO4 lot# <u>V46648</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	<u>22 97 E</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>22 84 C (5/17/02)</u>											
LPC												
ENDRN/DDT												

Analyst A.D.
revision 4/12/01

Verified by _____

Reviewed by DW
Date 6/4/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\052202\DFTPP.D

Acq On : 22 May 2002 12:09 pm

Sample :

Misc :

MS Integration Params: EVENTS.E

Vial: 1

Operator: Mclean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\052202.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3597

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	41.5	147584	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	35.7	127016	PASS
70	69	0.00	2	0.3	433	PASS
127	198	40	60	48.8	173440	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	355392	PASS
199	198	5	9	6.6	23504	PASS
275	198	10	30	20.4	72648	PASS
365	198	1	99	2.6	9315	PASS
441	443	1	100	75.9	40856	PASS
442	198	40	110	80.4	285568	PASS
443	442	17	23	18.8	53824	PASS

DFTPP.D 052202.M

Tue May 28 12:00:14 2002