

User: Vinci, David L
Westchester Dept. of Labs & Research
Date: 06/06/2002 Time: 15:50:48

Sample: AE12155

Analysis: GCMS GCMS Scan For Unknowns

Units: ug

Started: 6/6/02 15:50 Ended: 6/6/02 15:50 Analyst: DLV

Page: 1

	Component Name	Vol	Result	Quantity	NDI
1	Other unknown compounds		.		
2	Surr_2,4-DDD		89		10.0

Comments for Sample AE12155 - Analysis \$GCMS

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 06-06-2002 Time: 15:51:16

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recovery for 1 of the 43 compounds in the LCS Standard was low.

Quantitation Report

(QT Reviewed)

Data File : C:\HPCHEM\1\DATA\MAY2402\12155.D
 Acq On : 24 May 2002 11:26 pm
 Sample : gc/ms scan 5/22
 Misc :
 MS Integration Params: GOOFY.P
 Quant Time: May 28 11:58 2002

Vial: 9
 Operator: Morgan
 Inst : 209
 Multiplr: 1.00

Quant Results File: GCMS0520.RES

Quant Method : C:\HPCHEM\1\METHODS\GCMS0520.M (RTE Integrator)
 Title : 209 625/TCLP calibration table
 Last Update : Tue May 28 11:39:13 2002
 Response via : Initial Calibration
 DataAcq Meth : GCMS0520

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	12.14	152	521589	40.00	ug/l	0.00
10) Naphthalene-d8	15.77	136	2052162	40.00	ug/l	-0.02
17) Acenaphthene-d10	21.08	164	975404	40.00	ug/l	0.00
30) Phenanthrene-d10	25.52	188	1382301	40.00	ug/l	-0.01
38) Chrysene-d12	33.57	240	664208	40.00	ug/l	-0.03
44) Perylene-d12	37.78	264	428914	40.00	ug/l	-0.02

System Monitoring Compounds

28) TCMX	23.22	209	173680	35.43	ug/L	-0.01
Spiked Amount	40.000		Recovery	=	88.58%	

Target Compounds

					Qvalue
2) N-nitrosodimethylamine	0.00	74	0	N.D.	
3) bis(2-Chloroethyl) ether	0.00	93	0	N.D.	
4) 1,3-Dichlorobenzene	12.15	146	604	N.D.	
5) 1,4-Dichlorobenzene	12.15	146	604	N.D.	
6) 1,2-Dichlorobenzene	0.00	146	0	N.D.	
7) 2,2'-oxybis(1-Chloropropan	0.00	45	0	N.D.	
8) N-Nitrosodi-n-propylamine	0.00	70	0	N.D.	
9) Hexachloroethane	0.00	117	0	N.D.	
11) Nitrobenzene	0.00	77	0	N.D.	
12) Isophorone	0.00	82	0	N.D.	
13) bis(2-Chloroethoxy) methane	0.00	93	0	N.D.	
14) 1,2,4-Trichlorobenzene	0.00	180	0	N.D.	
15) Naphthalene	15.83	128	276	N.D.	
16) Hexachlorobutadiene	0.00	225	0	N.D.	
18) Hexachlorocyclopentadiene	0.00	237	0	N.D.	
19) 2-Chloronaphthalene	0.00	162	0	N.D.	
20) Dimethylphthalate	0.00	163	0	N.D.	
21) 2,6-Dinitrotoluene	0.00	165	0	N.D.	
22) Acenaphthylene	0.00	152	0	N.D.	
23) Acenaphthene	0.00	153	0	N.D.	
24) 2,4-Dinitrotoluene	0.00	165	0	N.D.	
25) Diethylphthalate	22.57	149	450	N.D.	
26) 4-Chlorophenylphenylether	23.22	204	1104	N.D.	
27) Fluorene	0.00	166	0	N.D.	
29) Azobenzene	0.00	77	0	N.D.	
31) N-Nitrosodiphenylamine	23.24	168	280	N.D.	
32) 4-Bromophenylphenylether	0.00	248	0	N.D.	
33) Hexachlorobenzene	0.00	284	0	N.D.	
34) Phenanthrene	25.52	178	249	N.D.	

(#) = qualifier out of range (m) = manual integration

12155.D GCMS0520.M Tue May 28 11:58:50 2002

Page 1

Quantitation Report

(QT Reviewed)

Data File : C:\HPCHEM\1\DATA\MAY2402\12155.D
Acq On : 24 May 2002 11:26 pm
Sample : gc/ms scan 5/22
Misc :
MS Integration Params: GOOFY.P
Quant Time: May 28 11:58 2002

Vial: 9
Operator: Morgan
Inst : 209
Multiplr: 1.00

Quant Results File: GCMS0520.RES

Quant Method : C:\HPCHEM\1\METHODS\GCMS0520.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Last Update : Tue May 28 11:39:13 2002
Response via : Initial Calibration
DataAcq Meth : GCMS0520

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
35) Anthracene	0.00	178	0		N.D.	
36) Di-n-butylphthalate	27.50	149	2105		N.D.	
37) Fluoranthene	0.00	202	0		N.D.	
39) Pyrene	0.00	202	0		N.D.	
40) Butylbenzylphthalate	0.00	149	0		N.D.	
41) Benzo(a)anthracene	33.57	228	1550		N.D.	
42) Bis(2-ethylhexyl)phthalate	33.80	149	299		N.D.	
43) Chrysene	33.57	228	1550		N.D.	
45) Di-n-Octylphthalate	0.00	149	0		N.D.	
46) Benzo(b)fluoranthene	0.00	252	0		N.D.	
47) Benzo(k)fluoranthene	0.00	252	0		N.D.	
48) Benzo(a)pyrene	37.78	252	1655		N.D.	
49) Indeno(1,2,3-cd)pyrene	0.00	276	0		N.D.	
50) Dibenz(a,h)anthracene	0.00	278	0		N.D.	
51) Benzo(g,h,i)perylene	0.00	276	0		N.D.	

(#) = qualifier out of range (m) = manual integration

12155.D GCMS0520.M Tue May 28 11:58:51 2002

Page 2


Quantitation Report

Data File : C:\HPCHEM\1\DATA\MAY2402\12155.D
 Acq On : 24 May 2002 11:26 pm
 Sample : gc/ms scan 5/22
 Misc :
 MS Integration Params: GOOFY.P
 Quant Time: May 28 11:58 2002

Vial: 9
 Operator: Morgan
 Inst : 209
 Multiplr: 1.00

Quant Results File: GCMS0520.RES

Method : C:\HPCHEM\1\METHODS\GCMS0520.M (RTE Integrator)
 Title : 209 625/TCLP calibration table
 Last Update : Tue May 28 11:39:13 2002
 Response via : Initial Calibration



LSC Area Percent Report

Data File : C:\HPCHEM\1\DATA\MAY2402\12155.D

Acq On : 24 May 2002 11:26 pm

Sample : gc/ms scan 5/22

Misc :

MS Integration Params: LSCINT.P

Vial: 9

Operator: Morgan

Inst : 209

Multiplr: 1.00

Method : C:\HPCHEM\1\METHODS\GCMS0520.M (RTE Integrator)

Title : 209 625/TCLP calibration table

Smoothing : OFF

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 1 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	peak area	peak % max.	% of total
1	8.118	331	335	363	rBV	283491	809738	20.99%	4.013%
2	12.139	768	773	794	rBV	1054704	2843181	73.69%	14.092%
3	15.774	1163	1169	1183	rBV	1812145	3782885	98.04%	18.749%
4	21.080	1741	1747	1760	rBV	1873380	3858328	100.00%	19.123%
5	23.219	1972	1980	1989	rBV	795995	1639652	42.50%	8.127%
6	25.524	2224	2231	2243	rBV2	1624750	3549618	92.00%	17.593%
7	33.575	3101	3108	3126	rBV2	940097	2104576	54.55%	10.431%
8	37.779	3558	3566	3581	rBV2	538650	1588480	41.17%	7.873%


Sum of corrected areas: 20176458

12155.D GCMS0520.M

Tue May 28 12:04:29 2002

LSC Report - Integrated Chromatogram

File : C:\HPCHEM\1\DATA\MAY2402\12155.D
 Operator : Morgan
 Acquired : 24 May 2002 11:26 pm using AcqMethod GCMS0520
 Instrument : 209
 Sample Name: gc/ms .scan 5/22
 Misc Info :
 Vial Number: 9
 Quant File :GCMS0520.RES (RTE Integrator)



Library Search Compound Report

Data File : C:\HPCHEM\1\DATA\MAY2402\12155.D
 Acq On : 24 May 2002 11:26 pm
 Sample : gc/ms scan 5/22
 Misc :
 MS Integration Params: LSCINT.P










Vial: 9
 Operator: Morgan
 Inst : 209
 Multiplr: 1.00

Quant Method : C:\HPCHEM\1\METHODS\GCMS0520.M (RTE Integrator)
 Title : 209 625/TCLP calibration table
 Library : C:\DATABASE\NBS75K.L

 Peak Number 1 2-Pentanone, 4-hydroxy-4-methy Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
8.12	11.39 ug/l	809738	1,4-Dichlorobenzene-d4	12.14

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	74
2	Acetone	58	C3H6O	000067-64-1	16
3	Formamide, N,N-diethyl-	101	C5H11NO	000617-84-5	12
4	Propanenitrile, 3-ethoxy-	99	C5H9NO	002141-62-0	9



Tentatively Identified Compound (LSC) summary

Operator ID: Morgan Date Acquired: 24 May 2002 11:26 pm
 Data File: C:\HPCHEM\1\DATA\MAY2402\12155.D
 Name: gc/ms scan 5/22
 Misc:
 Method: C:\HPCHEM\1\METHODS\GCMS0520.M (RTE Integrator)
 Title: 209 625/TCLP calibration table
 Library Searched: C:\DATABASE\NBS75K.L

TIC Top Hit name	RT	EstConc	Units	Area	IntStd	ISRT	ISArea	ISConc
2-Pentanone, 4-hydro	8.12	11.4	ug/l	809738	ISTD01	12.14	2843180	40.0

12155.D GCMS0520.M Tue May 28 12:04:31 2002