

Method 58mod

Date Extracted 5-22-02

Page 138

Ref

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spr add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1ml			1ml	1A	7			
REF					1ml			2B				
REF					↓			3A				
119026								2B				
11927								2A				
11928								8B				
11929	FB							9B				
11930								10B				
11931								9A				
11932								7A				
11933								8A				
12092								3B				
12161								3A				
12162								10B				
Remarks												
Reviewed by <u>C/B 8/23/02</u>												
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate - Extraction type - <u>sep fw</u> Solv. lot# <u>X070E07</u> NaCl lot# <u>V39618</u> Na2SO4 lot# <u>V46648</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	<u>2297E</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>TCMX 2294C</u>											
LPC												
ENDRIN/DDT												

low
sur
do
5/30

Analyst E.H.
revision 4/12/01

Verified by _____

Reviewed by DW 6/5/02
Date

Sequence Name: C:\MSDCHEM\1\SEQUENCE\052302.S

Comment:

Operator: Mclean

Data Path: C:\MSDCHEM\1\DATA\052302\

Pre-Seq Cmd:

Post-Seq Cmd:

Method Sections To Run	On A Barcode Mismatch
(X) Full Method	(X) Inject Anyway
() Reprocessing Only	() Don't Inject

Line	Type	Vial	DataFile	Method	Sample Name
------	------	------	----------	--------	-------------

1	Sample	1	DFTPP	508SCAN	
2	Sample	2	50CALCC	508SCAN	
3	Sample	3	522LB	508SCAN	5/22ad
4	Sample	4	522REF1	508SCAN	5/22ad
5	Sample	5	522REF2	508SCAN	5/22ad
6	Sample	6	11926	508SCAN	5/22ad
7	Sample	7	11927	508SCAN	5/22ad
8	Sample	8	11928	508SCAN	5/22ad
9	Sample	9	11929	508SCAN	5/22ad
10	Sample	10	11930	508SCAN	5/22ad
11	Sample	11	11931	508SCAN	5/22ad
12	Sample	12	11932	508SCAN	5/22ad
13	Sample	13	12092	508SCAN	5/22ad
14	Sample	14	12161	508SCAN	5/22ad
15	Sample	15	12162	508SCAN	5/22ad
16	Sample	1	DFTPB	508SCAN	
17	Sample	2	CAL10	508SCAN	

DFTPP

Data File : C:\MSDCHEM\1\DATA\052302\DFTPP.D

Acq On : 23 May 2002 9:05 am

Sample :

Misc :

MS Integration Params: EVENTS.E

Vial: 1

Operator: Mclean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\052202.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3596

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	43.8	129672	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.0	112504	PASS
70	69	0.00	2	0.6	629	PASS
127	198	40	60	50.6	149696	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	295744	PASS
199	198	5	9	6.6	19664	PASS
275	198	10	30	20.4	60296	PASS
365	198	1	99	2.5	7248	PASS
441	443	1	100	74.5	30400	PASS
442	198	40	110	68.6	202880	PASS
443	442	17	23	20.1	40800	PASS

DFTPP.D 052202.M

Wed May 29 09:36:30 2002