

Sequence Name: C:\MSDCHEM\1\SEQUENCE\052802.S

Comment:

Operator: McLean

Data Path: C:\MSDCHEM\1\DATA\052802\

Pre-Seq Cmd:

Post-Seq Cmd:

Method Sections To Run

On A Barcode Mismatch

☒ Full Method☒ Inject Anyway☐ Reprocessing Only☐ Don't Inject

Line	Type	Vial	DataFile	Method	Sample Name
1	Sample	1	DFTPP	508SCAN	
2	Sample	2	CAL50	508SCAN	
3	Sample	3	524LB	508SCAN	5/24 ad
4	Sample	4	524REF1	508SCAN	5/24 ad
5	Sample	5	524REF2	508SCAN	5/24 ad
6	Sample	6	12287	508SCAN	5/24 ad
7	Sample	7	12327	508SCAN	5/24 ad
8	Sample	8	11602	508SCAN	5/24 ad re-ext
9	Sample	9	11603	508SCAN	5/24 ad re-ext
10	Sample	10	12472	508SCAN	5/24 ad
11	Sample	11	12473	508SCAN	5/24 ad
12	Sample	12	12474	508SCAN	5/24 ad
13	Sample	13	12475	508SCAN	5/24 ad
14	Sample	1	DFTPB	508SCAN	5/24 ad
15	Sample	2	50CAL	508SCAN	5/24 ad

DFTPP

Data File : C:\MSDCHEM\1\DATA\052802\DFTPP.D
 Acq On : 28 May 2002 11:57 am
 Sample :
 Misc :
 MS Integration Params: EVENTS.E

Vial: 1
 Operator: McLean
 Inst : Instrumen
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\052202.M (Chemstation Integrator)
 Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3595

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	45.1	118344	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.0	97112	PASS
70	69	0.00	2	0.6	587	PASS
127	198	40	60	48.6	127312	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	262144	PASS
199	198	5	9	6.9	18016	PASS
275	198	10	30	19.5	51192	PASS
365	198	1	99	2.4	6176	PASS
441	443	1	100	78.2	23176	PASS
442	198	40	110	56.9	149056	PASS
443	442	17	23	19.9	29640	PASS

DFTPP.D 052202.M

Thu May 30 10:34:55 2002