

Ion 209.00 (208.70 to 209.70): TCMX.D

1	20.374 VV	0.036	93158	20.307	20.383
2	20.504 VB	0.033	2616417	20.383	20.573

7

DFTPP

Data File : C:\MSDCHEM\1\DATA\052902\DFTPP.D
Acq On : 29 May 2002 9:01 am
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: McLean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\052202.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3595

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	44.6	129392	PASS
68	69	0.00	2	2.0	2164	PASS
69	198	0.00	100	37.6	109168	PASS
70	69	0.00	2	0.7	727	PASS
127	198	40	60	49.0	142080	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	289984	PASS
199	198	5	9	6.6	19160	PASS
275	198	10	30	19.4	56296	PASS
365	198	1	99	2.2	6521	PASS
441	443	1	100	79.6	26672	PASS
442	198	40	110	59.5	172672	PASS
443	442	17	23	19.4	33488	PASS

DFTPP.D 052202.M

Mon Jun 03 09:25:59 2002