

Method 508 MOD.Date Extracted 5/28/02Page 2

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count
L.B.		500ml		1ml			1ml	6A	7	
Ref 1					1ml			5B		
Ref 2					1ml			10B		
12326								3A		
12476								9B		
12477								5A		
12478								10A		
12479	FB.							4B		
12531								9A		
12532								4A		
12533								6B		
12534								3B		
Remarks										
Reviewed by, <u>CB 8/23/02</u>										
drip rate must be between 5-15 ml/min for liq-liq extractors										
drip rate -										
Extraction type - <u>SEP. FOM.</u>		Solv. lot# <u>X05E10</u>		NaCl lot# <u>V</u>		Na2SO4 lot# <u>V46648</u>				
STANDARD	INT STD LOG#									
CALIBRATION										
REF SPIKE	<u>2297E</u>									
MDL/MDRF										
Q. CHECK										
INT. STD.										
SURROGATE	<u>22842</u>	<u>5/21/02</u>								
LPC										
ENDRINDOT										

Analyst A.D
Revision 4/12/01

Verified by _____

Reviewed by DN
Date 6/11/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\052902\DFTPP.D

Acq On : 29 May 2002 9:01 am

Sample :

Misc :

MS Integration Params: EVENTS.E

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\052202.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3595

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	44.6	129392	PASS
68	69	0.00	2	2.0	2164	PASS
69	198	0.00	100	37.6	109168	PASS
70	69	0.00	2	0.7	727	PASS
127	198	40	60	49.0	142080	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	289984	PASS
199	198	5	9	6.6	19160	PASS
275	198	10	30	19.4	56296	PASS
365	198	1	99	2.2	6521	PASS
441	443	1	100	79.6	26672	PASS
442	198	40	110	59.5	172672	PASS
443	442	17	23	19.4	33488	PASS

DFTPP.D 052202.M

Mon Jun 03 11:26:43 2002