

508 MOD.

Date Extracted

6/4/02

Page

10

Reviewed by,
JB 8/23/02

Reviewed by

Date _____

Abstract

A.P.

Revision 4/12/01

Verified by

DFTPP

Data File : C:\MSDCHEM\1\DATA\060402\DFTPP.D

Acq On : 4 Jun 2002 3:52 pm

Sample :

Misc :

MS Integration Params: EVENTS.E

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\060302.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3595

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	46.1	145088	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.2	120048	PASS
70	69	0.00	2	0.5	555	PASS
127	198	40	60	49.9	157120	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	314624	PASS
199	198	5	9	6.7	21096	PASS
275	198	10	30	19.3	60872	PASS
365	198	1	99	2.3	7286	PASS
441	443	1	100	72.5	28936	PASS
442	198	40	110	64.6	203328	PASS
443	442	17	23	19.6	39888	PASS

DFTPP.D 060302.M

Wed Jun 05 12:11:16 2002