

Method 508 mod Date Extracted 5/8/02 - and 4/26/02 Page 109

RE-Extract

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1ml			1ml	1B	7			
REF					1ml			1A				
REF					↓			2A				
9151	Q							3B				
9154	Q							9B				
9155								3A				
9331								9A				
9332								10B				
9568	FB							10A				
9569								2B				
Remarks												
drip rate must be between 5-15 ml/min for liq-liq extractors												
drip rate -												
Extraction type <u>SEP FUR</u>		Solv. lot# <u>X05E10</u>		NaCl lot# <u>V39614</u>		Na2SO4 lot# <u>V46648</u>						
STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	<u>2279E</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>GCMS</u>	<u>TCMX</u>	<u>22849</u>	<u>5/7/02</u>								
LPC												
ENDRIN/DDT												

Analyst RH
revision 4/12/01

Verified by _____

Reviewed by DR
Date 5/15/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\050902\DFTPA.D
 Acq On : 9 May 2002 3:17 pm
 Sample :
 Misc :
 MS Integration Params: EVENTS.E

Vial: 1
 Operator: Mclean
 Inst : Instrumen
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
 Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3603

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	46.9	163136	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.2	139648	PASS
70	69	0.00	2	0.5	764	PASS
127	198	40	60	52.6	182848	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	347648	PASS
199	198	5	9	6.7	23456	PASS
275	198	10	30	20.1	69752	PASS
365	198	1	99	2.6	8962	PASS
441	443	1	100	76.3	37904	PASS
442	198	40	110	75.3	261632	PASS
443	442	17	23	19.0	49656	PASS

DFTPA.D 050102.M Mon May 13 11:56:07 2002