

NYC-WTC_000022962

Sequence Name: C:\MSDCHEM\1\SEQUENCE\050902.S

Comment:

Operator: Mclean

Data Path: C:\MSDCHEM\1\DATA\050902\

Pre-Seq Cmd:

Post-Seq Cmd:

Method Sections To Run On A Barcode Mismatch

(X) Full Method (X) Inject Anyway

() Reprocessing Only () Don't Inject

Line	Type	Vial	DataFile	Method	Sample Name
------	------	------	----------	--------	-------------

1	Sample	1	DFTPA	508SCAN	
2	Sample	2	CAL20	508SCAN	
3	Sample	3	MC	508SCAN	
4	Sample	4	0508LB	508SCAN	5/8 kb
5	Sample	5	0508REF1	508SCAN	5/8 kb
6	Sample	6	0508REF2	508SCAN	5/8 kb
7	Sample	7	9570	508SCAN	5/8 kb
8	Sample	8	9571	508SCAN	5/8 kb
9	Sample	9	9572	508SCAN	5/8 kb
10	Sample	10	9574	508SCAN	5/8 kb
11	Sample	11	9575	508SCAN	5/8 kb
12	Sample	12	9576	508SCAN	5/8 kb
13	Sample	13	9577	508SCAN	5/8 kb
14	Sample	1	DFTPB	508SCAN	
15	Sample	14	CAL50	508SCAN	
16	Sample	15	508LB	508SCAN	5/8 eh
17	Sample	16	508REF1	508SCAN	5/8 eh
18	Sample	17	508REF2	508SCAN	5/8 eh
19	Sample	18	9151	508SCAN	5/8 eh
20	Sample	19	9154	508SCAN	5/8 eh
21	Sample	20	9155	508SCAN	5/8 eh
22	Sample	21	9331	508SCAN	5/8 eh
23	Sample	22	9332	508SCAN	5/8 eh
24	Sample	23	9568	508SCAN	5/8 eh
25	Sample	24	9569	508SCAN	5/8 eh

DFTPP

Data File : C:\MSDCHEM\1\DATA\050902\DFTPB.D

Vial: 1

Acq On : 10 May 2002 1:53 am

Operator: Mclean

Sample :

Inst : Instrumen

Misc :

Multiplr: 1.00

MS Integration Params: EVENTS.E

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3603

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	44.7	155648	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.5	134208	PASS
70	69	0.00	2	0.6	847	PASS
127	198	40	60	49.9	174016	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	348416	PASS
199	198	5	9	6.5	22672	PASS
275	198	10	30	20.7	72056	PASS
365	198	1	99	2.4	8530	PASS
441	443	1	100	74.7	38224	PASS
442	198	40	110	77.8	270912	PASS
443	442	17	23	18.9	51184	PASS

DFTPB.D 050102.M

Mon May 13 11:56:22 2002