

DFTPP

Data File : C:\MSDCHEM\1\DATA\051302\DFTPP.D
Acq On : 13 May 2002 9:37 am
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	46.2	138688	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.2	117664	PASS
70	69	0.00	2	0.5	646	PASS
127	198	40	60	49.9	149632	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	299968	PASS
199	198	5	9	6.8	20496	PASS
275	198	10	30	19.5	58472	PASS
365	198	1	99	2.2	6565	PASS
441	443	1	100	75.2	29920	PASS
442	198	40	110	67.9	203648	PASS
443	442	17	23	19.5	39776	PASS

DFTPP.D 050102.M

Tue May 14 11:50:09 2002

DFTPP

Data File : C:\MSDCHEM\1\DATA\051302\DFTPB.D
 Acq On : 13 May 2002 11:44 pm
 Sample :
 Misc :
 MS Integration Params: EVENTS.E

Vial: 1
 Operator: Mclean
 Inst : Instrumen
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
 Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	45.3	142272	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.0	122576	PASS
70	69	0.00	2	0.6	769	PASS
127	198	40	60	50.2	157696	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	314240	PASS
199	198	5	9	6.8	21360	PASS
275	198	10	30	20.0	62872	PASS
365	198	1	99	2.4	7400	PASS
441	443	1	100	71.6	27728	PASS
442	198	40	110	62.0	194944	PASS
443	442	17	23	19.9	38744	PASS

DFTPB.D 050102.M

Tue May 14 11:50:24 2002