

Method 508 mod

Date Extracted 5/10/02

Page 117

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1ml	1ml		1ml	6A	7			
Ref 1								6B				
Ref 2								8B				
10205	FB							8A				
10206								7A				
10895								7B				
10896								6A				
10897	FB							7A				
10898								8B				
10899								7B				
10900								6B				
Remarks												
Reviewed by, <i>[Signature]</i> 8/22/02												

rip rate must be between 5-15 ml/min for liq-liq extractors drip rate -

Extraction type sep fur Solv. lot# X05E10 NaCl lot# V39614 Na2SO4 lot# V46648

STANDARD	INT STD LOG#
CALIBRATION	
REF SPIKE	22795
MDL/MDRF	
Q. CHECK	
INT. STD.	
SURROGATE	TCMX 22849 5/7/02
LPC	
ENDRIN/DDT	

Analyst EH
Revision 4/12/01

Verified by _____

Reviewed by [Signature]
Date 5/21/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\051602\DFTPP.D
Acq On : 16 May 2002 2:32 pm
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3601

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	48.5	118736	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.6	99408	PASS
70	69	0.00	2	0.5	483	PASS
127	198	40	60	51.7	126424	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	244608	PASS
199	198	5	9	7.3	17944	PASS
275	198	10	30	18.9	46256	PASS
365	198	1	99	2.5	6230	PASS
441	443	1	100	75.3	20328	PASS
442	198	40	110	56.4	138048	PASS
443	442	17	23	19.5	26984	PASS

DFTPP.D 050102.M

Fri May 17 15:14:01 2002