


Method 508 MOD.

Date Extracted 5/15/02

Page 126

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
L.B.		500ml		1ml			1ml	4B	7			
Ref 1					1ml			3A				
Ref 2					1ml			4A				
11379								5A				
11380								1A				
11381								10A				
11382								9A				
11383								10B				
11384	F.B.							4B				
11409								3B				
Remarks												
Reviewed by, 												
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate - Extraction type - <u>Sep. Fur.</u> Solv. lot# <u>XO 7E07</u> NaCl lot# <u>V39614</u> Na2SO4 lot# <u>V46648</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF/SPIKE	2297E											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	TCHX 22846											
LPC												
ENDRIN/DDT												

Analyst A.D.
 Revision 4/12/01

Verified by _____

Reviewed by _____

DFTPP

Data File : C:\MSDCHEM\1\DATA\051702\DFTPP.D

Acq On : 17 May 2002 9:45 am

Sample :

Misc :

MS Integration Params: EVENTS.E

Vial: 1

Operator: Mclean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3604

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	41.9	142400	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.4	127120	PASS
70	69	0.00	2	0.5	627	PASS
127	198	40	60	49.3	167424	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	339648	PASS
199	198	5	9	6.7	22696	PASS
275	198	10	30	21.7	73664	PASS
365	198	1	99	2.8	9582	PASS
441	443	1	100	77.2	43656	PASS
442	198	40	110	87.0	295360	PASS
443	442	17	23	19.1	56560	PASS

DFTPP.D 050102.M

Tue May 21 12:01:21 2002