

DIFF

Data File : C:\MSDCHEM\1\DATA\051702\DFTPB.D

Acq On : 18 May 2002 12:11 am

Sample :

Misc :

MS Integration Params: EVENTS.E

Vial: 1

Operator: Mclean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)

Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3605

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	36.9	92176	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	32.2	80376	PASS
70	69	0.00	2	0.5	430	PASS
127	198	40	60	43.5	108600	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	249536	PASS
199	198	5	9	6.2	15572	PASS
275	198	10	30	21.8	54360	PASS
365	198	1	99	2.9	7117	PASS
441	443	1	100	86.8	35384	PASS
442	198	40	110	92.0	229632	PASS
443	442	17	23	17.8	40768	PASS

DFTPB.D 050102.M

Tue May 21 16:00:41 2002