

Method 508 modDate Extracted 5-16-02Page 129

Ref

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1ml			1ml	10B	7			
Ref 1					1ml			9A				
Ref 2					↓			9B				
11417								9A				
11418								10B				
11419								9B				
11420								10A				
11421								7B				
11422								10A				
11423		↓		↓			↓	10A	↓			
Remarks												
Reviewed by, <u>[Signature]</u>												
drip rate must be between 5-15 ml/min for liq-liq extractors												
drip rate -												
Extraction type <u>Sep Fun</u>		Solv. lot# <u>X07E01</u>		NaCl lot# <u>V39614</u>		Na2SO4 lot# <u>V46648</u>						
STANDARD	INT STD LOG#											
CALIBRATION												
REF/SPIKE	<u>2297E</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>TCMX</u>	<u>22846</u>		<u>5/10/02</u>								
LPC												
ENDRIN/DDT												

Analyst E. H.
revision 4/12/01

Verified by _____

Reviewed by [Signature]
Date 5/30/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\052002\DFTPP.D
Acq On : 20 May 2002 9:33 am
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\052202.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3604

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	39.3	101280	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.4	86104	PASS
70	69	0.00	2	0.4	387	PASS
127	198	40	60	45.9	118424	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	257856	PASS
199	198	5	9	6.3	16320	PASS
275	198	10	30	21.8	56240	PASS
365	198	1	99	2.8	7149	PASS
441	443	1	100	79.0	34152	PASS
442	198	40	110	89.1	229760	PASS
443	442	17	23	18.8	43216	PASS

DFTPP.D 052202.M

Wed May 22 15:54:30 2002