

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	10/2/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	9/28/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	10/2/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	10/2/2001	PAD
EPA94	Calcium	5500	ug/L	1000.0	10/2/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	10/2/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	10/2/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94	Iron	< MDL	ug/L	20.0	10/2/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	9/28/2001	PAD
SM18	Potassium	460	ug/L	100.0	10/3/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 1 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1200	ug/L	1000.0	10/2/2001	PAD
EPA94	Manganese	29	ug/L	20.0	10/2/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	10/2/2001	PAD
EPA94	Sodium	3600	ug/L	1000.0	10/2/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	10/2/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	10/1/2001	PAD
EPA79	Potable Metals Digestion	Completed			10/1/2001	PAD
SM18	Antimony	< MDL	ug/L	6	9/28/2001	PAD
SM18	Selenium	< MDL	ug/L	2	10/1/2001	PAD
SM18	Tin	< MDL	ug/L	100	10/5/2001	PAD
	Stat charge for HG TEST	Completed			10/2/2001 9:08 AM	PAD
SM18	Titanium	< MDL	ug/L	100	10/2/2001	PAD
EPA94	Thallium	< MDL	ug/L	2	10/3/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	10/2/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 2 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Zinc	< MDL	ug/L	20.0	10/2/2001	PAD
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	10/1/2001	DLV
	Aldrin	< MDL	ug/L	0.010	10/1/2001	DLV
	Chlordane	< MDL	ug/L	0.40	10/1/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	10/1/2001	DLV
	Endrin	< MDL	ug/L	0.020	10/1/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	10/1/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	10/1/2001	DLV
	Lindane	< MDL	ug/L	0.030	10/1/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	10/1/2001	DLV
	PCB's	< MDL	ug/L	0.20	10/1/2001	DLV
	Propachlor	< MDL	ug/L	0.040	10/1/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	10/1/2001	DLV
Volatile Organic Compounds						
524.2	*THM-Bromodichloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	*THM-Bromoform	< MDL	ug/L	2.0	9/28/2001	SV
524.2	*THM-Chloroform	< MDL	ug/L	.5	9/28/2001	SV
524.2	*THM-Dibromochloromethane	< MDL	ug/L	2.0	9/28/2001	SV
524.2	1,1,1,2-tetrachloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1,1- trichloroethane	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 3 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	1,1,2,2-tetrachloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1,2-trichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-dichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-Dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,3-trichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,3-trichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,4-trichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,4-trimethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3,5-trimethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,4-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	2,2-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	2-butanone (MEK)	< MDL	ug/L	2.0	9/28/2001	SV
524.2	2-chlorotoluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	4-chlorotoluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Benzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Bromobenzene	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 4 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	Bromochloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Bromomethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Carbon tetrachloride	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	cis-1,2-dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	cis-1,3-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Dibromomethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Dichlorodifluoromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Ethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Hexachlorobutadiene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Isopropylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Methyl iso-butyl ketone (MIBK)	< MDL	ug/L	1.0	9/28/2001	SV
524.2	Methyl tert-butyl ether (MTBE)	< MDL	ug/L	1.0	9/28/2001	SV
524.2	Methylene Chloride	< MDL	ug/L	.5	9/28/2001	SV
524.2	N-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	N-propylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Naphthalene	< MDL	ug/L	.5	9/28/2001	SV
524.2	O-xylene	< MDL	ug/L	.5	9/28/2001	SV
524.2	P & M-xylene	< MDL	ug/L	.5	9/28/2001	SV
524.2	P-isopropyltoluene	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 5 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	SEC-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Styrene	< MDL	ug/L	.5	9/28/2001	SV
524.2	TERT-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Tetrachloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Toluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	trans-1,2-dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	trans-1,3-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Trichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Trichlorofluoromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Vinyl chloride	< MDL	ug/L	.5	9/28/2001	SV
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	10/9/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	10/9/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	10/9/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	10/9/2001	DLV
	Alachlor	< MDL	ug/L	0.20	10/9/2001	DLV
	Atrazine	< MDL	ug/L	0.10	10/9/2001	DLV
	Benzo(a)pyrene	< MDL	ug/L	0.20	10/9/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	10/9/2001	DLV
	bis(2-Ethylhexyl)phthalate	0.77	ug/L	0.60	10/9/2001	DLV
	Butachlor	< MDL	ug/L	0.40	10/9/2001	DLV
	EPTC	< MDL	ug/L	1.0	10/9/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 6 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c132,k6491,7494,4017,p8869,88

Collected By: CURCIO

Comment:

Sample NYC DEP

19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CAT LEFF

ID of Source:

Collection 9/27/2001 AT 9:30:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23797

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Hexachlorobenzene	< MDL	ug/L	0.10	10/9/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	10/9/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	10/9/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	10/9/2001	DLV
	Molinate	< MDL	ug/L	0.90	10/9/2001	DLV
	Simazine	< MDL	ug/L	0.40	10/9/2001	DLV
	Terbacil	< MDL	ug/L	2.0	10/9/2001	DLV

Recoveries for three of the eighteen compounds in the MDL Standard and the LCS Standard were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges. The recovery for one of the three Surrogate Standards in this sample was above its acceptance range. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.76

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	11/1/2001	DLV
-------------------------	---	------	-----------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Environmental Laboratories

Date Approved: 11/5/2001

emailed

Original 11/5/2001

NYC ELAP # 10108

Page 7 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	10/2/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	9/28/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	10/2/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	10/2/2001	PAD
EPA94	Calcium	22000	ug/L	1000.0	10/2/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	10/2/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	10/2/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94	Iron	52	ug/L	20.0	10/2/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	9/28/2001	PAD
SM18	Potassium	1900	ug/L	100.0	10/3/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 1 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	7300	ug/L	1000.0	10/2/2001	PAD
EPA94	Manganese	82	ug/L	20.0	10/2/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	10/2/2001	PAD
EPA94	Sodium	27000	ug/L	1000.0	10/2/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	10/2/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	10/1/2001	PAD
EPA79	Potable Metals Digestion	Completed			10/1/2001	PAD
SM18	Antimony	< MDL	ug/L	6	9/28/2001	PAD
SM18	Selenium	< MDL	ug/L	2	10/1/2001	PAD
SM18	Tin	< MDL	ug/L	100	10/5/2001	PAD
	Stat charge for HG TEST	Completed			10/2/2001 9:11 AM	PAD
SM18	Titanium	< MDL	ug/L	100	10/2/2001	PAD
EPA94	Thallium	< MDL	ug/L	2	10/3/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	10/2/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 2 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Zinc	< MDL	ug/L	20.0	10/2/2001	PAD
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	10/1/2001	DLV
	Aldrin	< MDL	ug/L	0.010	10/1/2001	DLV
	Chlordane	< MDL	ug/L	0.40	10/1/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	10/1/2001	DLV
	Endrin	< MDL	ug/L	0.020	10/1/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	10/1/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	10/1/2001	DLV
	Lindane	< MDL	ug/L	0.030	10/1/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	10/1/2001	DLV
	PCB's	< MDL	ug/L	0.20	10/1/2001	DLV
	Propachlor	< MDL	ug/L	0.040	10/1/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	10/1/2001	DLV
Volatile Organic Compounds						
524.2	*THM-Bromodichloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	*THM-Bromoform	< MDL	ug/L	2.0	9/28/2001	SV
524.2	*THM-Chloroform	< MDL	ug/L	.5	9/28/2001	SV
524.2	*THM-Dibromochloromethane	< MDL	ug/L	2.0	9/28/2001	SV
524.2	1,1,1,2-tetrachloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1,1- trichloroethane	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 3 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	1,1,2,2-tetrachloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1,2-trichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-dichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-Dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,3-trichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,3-trichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,4-trichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,4-trimethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3,5-trimethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,4-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	2,2-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	2-butanone (MEK)	< MDL	ug/L	2.0	9/28/2001	SV
524.2	2-chlorotoluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	4-chlorotoluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Benzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Bromobenzene	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 4 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	Bromochloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Bromomethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Carbon tetrachloride	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	cis-1,2-dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	cis-1,3-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Dibromomethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Dichlorodifluoromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Ethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Hexachlorobutadiene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Isopropylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Methyl iso-butyl ketone (MIBK)	< MDL	ug/L	1.0	9/28/2001	SV
524.2	Methyl tert-butyl ether (MTBE)	< MDL	ug/L	1.0	9/28/2001	SV
524.2	Methylene Chloride	< MDL	ug/L	.5	9/28/2001	SV
524.2	N-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	N-propylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Naphthalene	< MDL	ug/L	.5	9/28/2001	SV
524.2	O-xylene	< MDL	ug/L	.5	9/28/2001	SV
524.2	P & M-xylene	< MDL	ug/L	.5	9/28/2001	SV
524.2	P-isopropyltoluene	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 5 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	SEC-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Styrene	< MDL	ug/L	.5	9/28/2001	SV
524.2	TERT-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Tetrachloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Toluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	trans-1,2-dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	trans-1,3-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Trichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Trichlorofluoromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Vinyl chloride	< MDL	ug/L	.5	9/28/2001	SV
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	10/10/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	10/10/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	10/10/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	10/10/2001	DLV
	Alachlor	< MDL	ug/L	0.20	10/10/2001	DLV
	Atrazine	< MDL	ug/L	0.10	10/10/2001	DLV
	Benzo(a)pyrene	< MDL	ug/L	0.20	10/10/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	10/10/2001	DLV
	bis(2-Ethylhexyl)phthalate	1.2	ug/L	0.60	10/10/2001	DLV
	Butachlor	< MDL	ug/L	0.40	10/10/2001	DLV
	EPTC	< MDL	ug/L	1.0	10/10/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 6 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c123,k7603,7398,7692,p8866,88

Collected By: CURCIO

Comment:

Sample CRO GH

ID of Source:

Collection 9/27/2001 AT 11:00:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23798

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Hexachlorobenzene	< MDL	ug/L	0.10	10/10/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	10/10/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	10/10/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	10/10/2001	DLV
	Molinate	< MDL	ug/L	0.90	10/10/2001	DLV
	Simazine	< MDL	ug/L	0.40	10/10/2001	DLV
	Terbacil	< MDL	ug/L	2.0	10/10/2001	DLV

The recoveries for three of the eighteen compounds in the LCS Standard were outside of their acceptance ranges. The level of bis(2-Ethylhexyl)phthalate was 1.4 ug/L in the Laboratory Blank.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	11/1/2001	DLV
-------------------------	---	------	-----------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds generally different than those present in the Laboratory Blank.

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYC ELAP # 10108

Page 7 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
EPA94	Silver	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	10/2/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	9/28/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	10/2/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	10/2/2001	PAD
EPA94	Calcium	5500	ug/L	1000.0	10/2/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	10/2/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	10/2/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	10/2/2001	PAD
EPA94	Iron	< MDL	ug/L	20.0	10/2/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	9/28/2001	PAD
SM18	Potassium	470	ug/L	100.0	10/3/2001	PAD

Approved Pam Diisizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 1 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1200	ug/L	1000.0	10/2/2001	PAD
EPA94	Manganese	26	ug/L	20.0	10/2/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	10/2/2001	PAD
EPA94	Sodium	3600	ug/L	1000.0	10/2/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	10/2/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	10/1/2001	PAD
EPA'79	Potable Metals Digestion	Completed			10/1/2001	PAD
SM18	Antimony	< MDL	ug/L	6	9/28/2001	PAD
SM18	Selenium	< MDL	ug/L	2	10/1/2001	PAD
SM18	Tin	< MDL	ug/L	100	10/5/2001	PAD
	Stat charge for HG TEST	Completed			10/2/2001 9:13 AM	PAD
SM18	Titanium	< MDL	ug/L	100	10/2/2001	PAD
EPA94	Thallium	< MDL	ug/L	2	10/3/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	10/2/2001	PAD

Approved Pam Dilisizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 2 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Zinc	< MDL	ug/L	20.0	10/2/2001	PAD
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	10/1/2001	DLV
	Aldrin	< MDL	ug/L	0.010	10/1/2001	DLV
	Chlordane	< MDL	ug/L	0.40	10/1/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	10/1/2001	DLV
	Endrin	< MDL	ug/L	0.020	10/1/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	10/1/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	10/1/2001	DLV
	Lindane	< MDL	ug/L	0.030	10/1/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	10/1/2001	DLV
	PCB's	< MDL	ug/L	0.20	10/1/2001	DLV
	Propachlor	< MDL	ug/L	0.040	10/1/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	10/1/2001	DLV
Volatile Organic Compounds						
524.2	*THM-Bromodichloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	*THM-Bromoform	< MDL	ug/L	2.0	9/28/2001	SV
524.2	*THM-Chloroform	< MDL	ug/L	.5	9/28/2001	SV
524.2	*THM-Dibromochloromethane	< MDL	ug/L	2.0	9/28/2001	SV
524.2	1,1,1,2-tetrachloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1,1- trichloroethane	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 3 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	1,1,2,2-tetrachloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1,2-trichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-dichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-Dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,1-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,3-trichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,3-trichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,4-trichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2,4-trimethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,2-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3,5-trimethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,3-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	1,4-dichlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	2,2-dichloropropane	< MDL	ug/L	.5	9/28/2001	SV
524.2	2-butanone (MEK)	< MDL	ug/L	2.0	9/28/2001	SV
524.2	2-chlorotoluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	4-chlorotoluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Benzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Bromobenzene	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 4 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	Bromochloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Bromomethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Carbon tetrachloride	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chlorobenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chloroethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Chloromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	cis-1,2-dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	cis-1,3-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Dibromomethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Dichlorodifluoromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Ethylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Hexachlorobutadiene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Isopropylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Methyl iso-butyl ketone (MIBK)	< MDL	ug/L	1.0	9/28/2001	SV
524.2	Methyl tert-butyl ether (MTBE)	< MDL	ug/L	1.0	9/28/2001	SV
524.2	Methylene Chloride	< MDL	ug/L	.5	9/28/2001	SV
524.2	N-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	N-propylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Naphthalene	< MDL	ug/L	.5	9/28/2001	SV
524.2	O-xylene	< MDL	ug/L	.5	9/28/2001	SV
524.2	P & M-xylene	< MDL	ug/L	.5	9/28/2001	SV
524.2	P-isopropyltoluene	< MDL	ug/L	.5	9/28/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 5 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	SEC-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Styrene	< MDL	ug/L	.5	9/28/2001	SV
524.2	TERT-butylbenzene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Tetrachloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Toluene	< MDL	ug/L	.5	9/28/2001	SV
524.2	trans-1,2-dichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	trans-1,3-dichloropropene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Trichloroethene	< MDL	ug/L	.5	9/28/2001	SV
524.2	Trichlorofluoromethane	< MDL	ug/L	.5	9/28/2001	SV
524.2	Vinyl chloride	< MDL	ug/L	.5	9/28/2001	SV
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	10/9/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	10/9/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	10/9/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	10/9/2001	DLV
	Alachlor	< MDL	ug/L	0.20	10/9/2001	DLV
	Atrazine	< MDL	ug/L	0.10	10/9/2001	DLV
	Benzo(a)pyrene	< MDL	ug/L	0.20	10/9/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	10/9/2001	DLV
	bis(2-Ethylhexyl)phthalate	< MDL	ug/L	0.60	10/9/2001	DLV
	Butachlor	< MDL	ug/L	0.40	10/9/2001	DLV
	EPTC	< MDL	ug/L	1.0	10/9/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYS ELAP # 10108

Page 6 of 7

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP
19 WEST LAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c172,k4904,3519,6956,4128,p88

Collected By: CURCIO

Comment:

Sample DEL 18

ID of Source:

Collection 9/27/2001 AT 9:45:00

Submitted 9/27/2001 AT 4:23:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD23799

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Hexachlorobenzene	< MDL	ug/L	0.10	10/9/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	10/9/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	10/9/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	10/9/2001	DLV
	Molinate	< MDL	ug/L	0.90	10/9/2001	DLV
	Simazine	< MDL	ug/L	0.40	10/9/2001	DLV
	Terbacil	< MDL	ug/L	2.0	10/9/2001	DLV

Recoveries for three of the eighteen compounds in the MDL Standard and the LCS Standard were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	11/1/2001	DLV
-------------------------	---	------	-----------	-----

A scan of the extract prepared for the 508 analysis, from 35 - 550 amu using a mass spectrometer, does not reveal the presence of any non-target compounds generally different than those present in the Laboratory Blank.

Approved Pam Dilsizian

QA Officer

Date Approved: 11/5/2001

Environmental Laboratories

emailed

Original 11/5/2001

NYC ELAP # 10108

Page 7 of 7