


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/31/2001 Time: 11:43:46

Sample: AD23947
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/11/01 11:36 Ended: 10/23/01 11:36 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.2		1.0
20	Surr_Triphenyl phosphate		5.1		1.0
21	Surr_Perylene-d12		3.4		1.0

Chromatogram Plot File: E:\23947E Date: Oct-24-1991 10:47:13
Comment: BOHLK; METHOD 525; INST 204; 10/23/01; SAMPLE 23947
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 7613281



Integration Report Quanfile: 23947E Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/23/01; SAMPLE 23947
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:28	748	Auto	2,212,841	1,274,397
2	phenanthrene d-10 (IS)	17:11	1031	Auto	3,498,884	1,551,948
3	chrysene d-12(IS)	22:53	1373	Auto	2,591,312	1,092,191
4	p-terphenyl d-14(IS)	20:59	1259	Auto	3,767,293	2,402,908
5	acenaphthene d-10(SURR)	12:28	748	Auto	2,212,841	1,274,397
6	phenanthrene d-10(SURR)	17:11	1031	Auto	3,498,884	1,551,948
7	chrysene d-12 (SURR)	22:53	1373	Auto	2,591,312	1,092,191
8	1,3-dimethyl-2-nitrobenzene	8:08	488	Auto	1,357,553	781,090
9	triphenyl phosphate (S)	22:16	1336	Auto	1,760,223	721,871
10	perylene d-12 (SURR)	27:23	1643	Auto	1,114,052	203,473
11	bis(2-ethylhexyl)adipate	22:16	1336	Auto	21,188	4,702
12	bis(2-ethylhexyl)phthalate	23:11	1391	Auto	382,072	94,492
13	2,6-dinitrotoluene	12:07	727	Auto	26,295	13,074
14	2,4-dinitrotoluene	13:15	795	Auto	6,046	3,567
15	4,4'-dichlorodiphenyl ether	20:55	1255	Auto	1,943	1,943

Quantitation Report Quanfile: 23947E Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/23/01; SAMPLE 23947
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	748	12:28	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1031	17:11	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1373	22:53	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1259	20:59	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	748	12:28	BB	4.060	UG/L
6	phenanthrene d-10(SURR)	A	1031	17:11	BV	4.433	UG/L
7	chrysene d-12 (SURR)	A	1373	22:53	BB	4.800	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	488	8:08	BV	5.227	UG/L
9	triphenyl phosphate (S)	A	1336	22:16	BB	5.135	UG/L
10	perylene d-12 (SURR)	A	1643	27:23	BB	3.414	UG/L
16	bis(2-ethylhexyl)adipate	A	1336	22:16	MM	0.040	UG/L
17	bis(2-ethylhexyl)phthalate	A	1391	23:11	VB	0.432	UG/L
20	2,6-dinitrotoluene	A	727	12:07	BB	0.182	UG/L
21	2,4-dinitrotoluene	A	795	13:15	BB	0.036	UG/L
28	4,4'-dichlorodiphenyl ether	A	1255	20:55	BB	0.006	UG/L

Reviewed IS/surrogates

Comments for Sample AD23947 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-03-2001 Time: 13:33:41

The recoveries for five of the eighteen compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample were below their acceptance ranges. The Spiked Duplicate Sample was lost due to a laboratory accident. The level of bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 24 ug/L due to a laboratory accident.