



User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/31/2001 Time: 11:48:55

Sample: AD23995
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/11/01 11:36 Ended: 10/23/01 11:36 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.3		1.0
20	Surr_Triphenyl phosphate		5.8		1.0
21	Surr_Perylene-d12		3.7		1.0



Chromatogram Plot File: E:\23995J Date: Oct-24-1991 15:39:11
Comment: BOHLK; METHOD 525; INST 204; 10/23/01; SAMPLE 23995
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 9521587



Integration Report Quanfile: 23995J Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/23/01; SAMPLE 23995
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:27	747	Auto	2,242,100	1,249,709
2	phenanthrene d-10 (IS)	17:10	1030	Auto	3,633,046	1,694,610
3	chrysene d-12 (IS)	22:52	1372	Auto	2,235,505	895,614
4	p-terphenyl d-14 (IS)	20:58	1258	Auto	3,806,689	2,491,489
5	acenaphthene d-10 (SURR)	12:27	747	Auto	2,242,100	1,249,709
6	phenanthrene d-10 (SURR)	17:10	1030	Auto	3,633,046	1,694,610
7	chrysene d-12 (SURR)	22:52	1372	Auto	2,235,505	895,614
8	1,3-dimethyl-2-nitrobenzene	8:07	487	Auto	1,402,406	779,133
9	triphenyl phosphate (S)	22:15	1335	Auto	1,721,180	724,444
10	perylene d-12 (SURR)	27:22	1642	Auto	1,029,267	187,906
11	bis(2-ethylhexyl)adipate	22:07	1327	Auto	33,895	8,440
12	bis(2-ethylhexyl)phthalate	23:10	1390	Auto	310,225	76,800
13	2,6-dinitrotoluene	12:07	727	Auto	77,697	38,408
14	2,4-dinitrotoluene	13:14	794	Auto	15,601	5,134
15	4,4'-dichlorodiphenyl ether	20:54	1254	Auto	8,901	3,926

Quantitation Report

Quanfile: 23995J

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 10/23/01; SAMPLE 23995

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	747	12:27	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1030	17:10	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1372	22:52	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1258	20:58	BV	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	747	12:27	BV	4.071	UG/L
6	phenanthrene d-10 (SURR)	A	1030	17:10	BV	4.555	UG/L
7	chrysene d-12 (SURR)	A	1372	22:52	BB	4.098	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	487	8:07	BV	5.330	UG/L
9	triphenyl phosphate (S)	A	1335	22:15	BB	5.820	UG/L
10	perylene d-12 (SURR)	A	1642	27:22	BV	3.656	UG/L
16	bis(2-ethylhexyl)adipate	A	1327	22:07	MM	0.073	UG/L
17	bis(2-ethylhexyl)phthalate	A	1390	23:10	BV	0.406	UG/L
20	2,6-dinitrotoluene	A	727	12:07	BV	0.526	UG/L
21	2,4-dinitrotoluene	A	794	13:14	BB	0.090	UG/L
28	4,4'-dichlorodiphenyl ether	A	1254	20:54	BV	0.026	UG/L

Reviewed IS / Surrogates

Comments for Sample AD23995 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-03-2001 Time: 14:05:25

The recoveries for five of the eighteen compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample were below their acceptance ranges. The Spiked Duplicate Sample was lost due to a laboratory accident. The level of bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 24 ug/L due to a laboratory accident. The recovery for one of the three Surrogate Standards in this sample was below its acceptance range.