Action Taken On Failed QC Analyses

LAB #	DATE EXT.	COMMENT:	NAME
24235-		525	LB
24291			
		calibrated / quantitated EPTC using total mass instead of mass 128	
	10/15/01		
		calibrated / quantitated terbacil using CAD stds C5 - C8.	
		There were some truncated adipate peaks (for mass 129)	
		Manually checked LBs + samples for adipate, EPTC and simazine	
		Sample 24290 was injected on 10/30, whereas sample 24290 MS + MSD were inj'd on 10/25/01.	