

AD 24220

Cancelled 525.2 test
however results are clear

DW

11/14/1

call 10/2/1

Batch A100301

Chromatogram Plot

File: F:\24220H

Date: Oct-26-1991 02:38:52

Comment: BOHLK; METHOD 525; INST 204; 10/25/01; SAMPLE 24220

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 6171502



Integration Report Quanfile: 24220H Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 10/25/01; SAMPLE 24220
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:28	748	Auto	495,334	251,320
2	phenanthrene d-10 (IS)	17:10	1030	Auto	825,020	311,180
3	chrysene d-12 (IS)	22:53	1373	Auto	646,069	221,593
4	p-terphenyl d-14 (IS)	20:58	1258	Auto	3,720,491	1,958,826
5	acenaphthene d-10 (SURR)	12:28	748	Auto	495,334	251,320
6	phenanthrene d-10 (SURR)	17:10	1030	Auto	825,020	311,180
7	chrysene d-12 (SURR)	22:53	1373	Auto	646,069	221,593
8	1,3-dimethyl-2-nitrobenzene	8:07	487	Auto	360,256	157,571
9	triphenyl phosphate (S)	22:15	1335	Auto	472,130	156,951
10	perylene d-12 (SURR)	27:24	1644	Auto	364,018	57,205
11	bis(2-ethylhexyl)adipate	22:07	1327	Auto	16,478	4,448
12	bis(2-ethylhexyl)phthalate	23:10	1390	Auto	151,281	43,785
13	benzo(a)pyrene	27:11	1631	Auto	9,851	1,847

Quantitation Report

Quanfile: 24220H

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 10/25/01; SAMPLE 24220

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	748	12:28	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1030	17:10	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1373	22:53	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1258	20:50	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	748	12:28	BB	0.921	UG/L
6	phenanthrene d-10 (SURR)	A	1030	17:10	BV	1.059	UG/L
7	chrysene d-12 (SURR)	A	1373	22:53	BB	1.212	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	487	8:07	BV	6.197	UG/L
9	triphenyl phosphate (S)	A	1335	22:15	BB	5.524	UG/L
10	perylene d-12 (SURR)	A	1644	27:24	BB	4.474	UG/L
16	bis(2-ethylhexyl)adipate	A	1327	22:07	MM	0.123	UG/L
17	bis(2-ethylhexyl)phthalate	A	1390	23:10	VB	0.690	UG/L
18	benzo(a)pyrene	A	1631	27:11	BB	0.078	UG/L

normal
I.S. counts
would lower
target analyte
results even
more.