


User: Azzo Murad, Nehla
Westchester Dept. of Labs & Research
Date: 11/07/2001 Time: 10:18:10

Sample: AD24289
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/15/01 10:08 Ended: 10/25/01 10:08 Analyst: NAM
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.2		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.6		1.0
21	Surr_Perylene-d12		4.0		1.0

Chromatogram Plot File: C:\24289M Date: Oct-26-1991 09:40:33
Comment: BOHLK; METHOD 525; INST 204; 10/25/01; SAMPLE 24289
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 11383750



Integration Report Quanfile: 24289M Quan Entries: 17
 Comment: BOHLK; METHOD 525; INST 204; 10/25/01; SAMPLE 24289
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:29	749	Auto	2,219,716	1,132,716
2	phenanthrene d-10 (IS)	17:11	1031	Auto	3,473,451	1,485,986
3	chrysene d-12 (IS)	22:54	1374	Auto	2,412,537	860,422
4	p-terphenyl d-14 (IS)	20:59	1259	Auto	3,986,148	2,403,667
5	acenaphthene d-10 (SURR)	12:29	749	Auto	2,219,716	1,132,716
6	phenanthrene d-10 (SURR)	17:11	1031	Auto	3,473,451	1,485,986
7	chrysene d-12 (SURR)	22:54	1374	Auto	2,412,537	860,422
8	1,3-dimethyl-2-nitrobenzene	8:08	488	Auto	1,324,690	704,517
9	triphenyl phosphate (S)	22:16	1336	Auto	1,779,406	685,477
10	perylene d-12 (SURR)	27:24	1644	Auto	1,230,312	234,043
11	hexachlorocyclopentadiene	10:19	619	Auto	1,333	1,333
12	bis(2-ethylhexyl)adipate	22:16	1336	Auto	27,113	6,923
13	bis(2-ethylhexyl)phthalate	23:11	1391	Auto	972,067	316,572
14	2,6-dinitrotoluene	12:08	728	Auto	53,450	27,289
15	2,4-dinitrotoluene	13:15	795	Auto	4,509	2,279
16	metribuzin	18:26	1106	Auto	2,141	1,081
17	4,4'-dichlorodiphenyl ether	20:55	1255	Auto	1,386	1,386

Quantitation Report

Quanfile: 24289M

Quan Entries: 17

Comment: BOHLK; METHOD 525; INST 204; 10/25/01; SAMPLE 24289

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	749	12:29	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1031	17:11	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1374	22:54	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1259	20:59	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	749	12:29	BV	3.849	UG/L
6	phenanthrene d-10(SURR	A	1031	17:11	BV	4.159	UG/L
7	chrysene d-12 (SURR)	A	1374	22:54	BB	4.223	UG/L
8	1,3-dimethyl-2-nitrobe	A	488	8:08	BV	5.085	UG/L
9	triphenyl phosphate (S	A	1336	22:16	BV	5.575	UG/L
10	perylene d-12 (SURR)	A	1644	27:24	BB	4.049	UG/L
11	hexachlorocyclopentadi	A	619	10:19	BB	0.014	UG/L
16	bis(2-ethylhexyl)adipa	A	1336	22:16	MM	0.054	UG/L
17	bis(2-ethylhexyl)phtha	A	1391	23:11	BB	1.202	UG/L
20	2,6-dinitrotoluene	A	728	12:08	BB	0.367	UG/L
21	2,4-dinitrotoluene	A	795	13:15	BB	0.027	UG/L
25	metribuzin	A	1106	18:26	BB	0.009	UG/L
28	4,4'-dde	A	1255	20:55	BB	0.005	UG/L

Reviewed 15/surrogate

Comments for Sample AD24289 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-15-2001 Time: 12:02:15

The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range. This is probably due to laboratory contamination.