


User: Bohik, Ted
Westchester Dept. of Labs & Research
Date: 11/13/2001 Time: 14:28:31

Sample: AD24333
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/16/01 14:24 Ended: 10/30/01 14:24 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.9		1.0
20	Surr_Triphenyl phosphate		5.7		1.0
21	Surr_Perylene-d12		4.3		1.0

Chromatogram Plot File: E:\24333 Date: Oct-30-1991 19:55:00
Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24333
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 4440827



Integration Report Quanfile: 24333 Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24333
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:25	745	Auto	1,562,327	914,318
2	phenanthrene d-10 (IS)	17:07	1027	Auto	2,481,241	1,091,030
3	chrysene d-12 (IS)	22:50	1370	Auto	1,712,136	658,109
4	p-terphenyl d-14 (IS)	20:56	1256	Auto	3,029,278	1,402,194
5	acenaphthene d-10 (SURR)	12:25	745	Auto	1,562,327	914,318
6	phenanthrene d-10 (SURR)	17:07	1027	Auto	2,481,241	1,091,030
7	chrysene d-12 (SURR)	22:50	1370	Auto	1,712,136	658,109
8	1,3-dimethyl-2-nitrobenzene	8:05	485	Auto	933,677	473,930
9	triphenyl phosphate (S)	22:13	1333	Auto	1,328,830	586,443
10	perylene d-12 (SURR)	27:17	1637	Auto	917,428	175,650
11	hexachlorocyclopentadiene	10:16	616	Auto	886	886
12	hexachlorobenzene	15:34	934	Auto	12,101	3,775
13	bis(2-ethylhexyl)adipate	22:13	1333	Auto	36,838	4,065
14	bis(2-ethylhexyl)phthalate	23:07	1387	Auto	242,462	59,620
15	2,6-dinitrotoluene	12:05	725	Auto	30,413	15,483
16	4,4'-dichlorodiphenyl ether	20:52	1252	Auto	8,756	6,072

renewed IS/surr's

Quantitation Report Quanfile: 24333 Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24333
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	745	12:25	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1027	17:07	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1370	22:50	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1256	20:56	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	745	12:25	BV	4.081	UG/L
6	phenanthrene d-10 (SURR)	A	1027	17:07	BV	4.419	UG/L
7	chrysene d-12 (SURR)	A	1370	22:50	BB	4.180	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	485	8:05	BV	4.888	UG/L
9	triphenyl phosphate (S)	A	1333	22:13	BB	5.695	UG/L
10	perylene d-12 (SURR)	A	1637	27:17	BB	4.261	UG/L
11	hexachlorocyclopentadiene	A	616	10:16	BB	0.014	UG/L
12	hexachlorobenzene	A	934	15:34	BB	0.058	UG/L
16	bis(2-ethylhexyl)adipate	A	1333	22:13	MM	0.118	UG/L
17	bis(2-ethylhexyl)phthalate	A	1387	23:07	VB	0.453	UG/L
20	2,6-dinitrotoluene	A	725	12:05	BB	0.311	UG/L
28	4,4'-dichlorodiphenyl ether	A	1252	20:52	BB	0.040	UG/L

Comments for Sample AD24333 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-21-2001 Time: 16:26:21

The recoveries for 4 of the 18 compounds in the MDL Standard and 2 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.