


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 11/13/2001 Time: 14:32:35

Sample: AD24397
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 10/16/01 14:24 Ended: 10/30/01 14:24 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.9		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21	Surr_Perylene-d12		4.3		1.0

Chromatogram Plot File: E:\24397 Date: Oct-30-1991 22:12:28
Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24397
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5118395



Integration Report Quanfile: 24397 Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24397
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:25	745	Auto	1,641,844	1,074,742
2	phenanthrene d-10 (IS)	17:07	1027	Auto	2,480,181	1,095,459
3	chrysene d-12(IS)	22:50	1370	Auto	1,731,065	645,003
4	p-terphenyl d-14(IS)	20:56	1256	Auto	2,794,970	1,597,364
5	acenaphthene d-10(SURR)	12:25	745	Auto	1,641,844	1,074,742
6	phenanthrene d-10(SURR)	17:07	1027	Auto	2,480,181	1,095,459
7	chrysene d-12 (SURR)	22:50	1370	Auto	1,731,065	645,003
8	1,3-dimethyl-2-nitrobenzene	8:05	485	Auto	978,081	519,602
9	triphenyl phosphate (S)	22:13	1333	Auto	1,291,468	585,601
10	perylene d-12 (SURR)	27:17	1637	Auto	939,997	172,129
11	hexachlorocyclopentadiene	10:16	616	Auto	1,462	1,462
12	bis(2-ethylhexyl)adipate	22:13	1333	Auto	27,051	5,158
13	bis(2-ethylhexyl)phthalate	23:07	1387	Auto	311,975	87,802
14	2,6-dinitrotoluene	12:04	724	Auto	29,359	14,541
15	2,4-dinitrotoluene	13:12	792	Auto	2,207	1,345

reviewed IS/surrogates

Quantitation Report Quanfile: 24397 Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24397
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	745	12:25	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1027	17:07	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1370	22:50	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1256	20:56	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	745	12:25	BB	4.648	UG/L
6	phenanthrene d-10 (SURR)	A	1027	17:07	BV	4.788	UG/L
7	chrysene d-12 (SURR)	A	1370	22:50	BB	4.580	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	485	8:05	BB	4.872	UG/L
9	triphenyl phosphate (S)	A	1333	22:13	BB	5.474	UG/L
10	perylene d-12 (SURR)	A	1637	27:17	BB	4.318	UG/L
11	hexachlorocyclopentadiene	A	616	10:16	BB	0.022	UG/L
16	bis(2-ethylhexyl)adipate	A	1333	22:13	MM	0.086	UG/L
17	bis(2-ethylhexyl)phthalate	A	1387	23:07	VV	0.576	UG/L
20	2,6-dinitrotoluene	A	724	12:04	BB	0.286	UG/L
21	2,4-dinitrotoluene	A	792	13:12	BB	0.020	UG/L

Comments for Sample AD24397 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-21-2001 Time: 16:45:49

The recoveries for 4 of the 18 compounds in the MDL Standard and 2 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.