


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/13/2001 Time: 15:07:00

Sample: AD24472
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/16/01 14:58 Ended: 10/30/01 14:58 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.2		1.0
20	Surr_Triphenyl phosphate		5.7		1.0
21	Surr_Perylene-d12		4.5		1.0

Chromatogram Plot File: E:\24472 Date: Oct-31-1991 07:56:35
Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24472
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 3910823



Integration Report Quanfile: 24472 Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24472
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:25	745	Auto	1,323,814	845,312
2	phenanthrene d-10 (IS)	17:07	1027	Auto	2,030,899	757,975
3	chrysene d-12 (IS)	22:49	1369	Auto	1,479,266	594,409
4	p-terphenyl d-14 (IS)	22:49	1369	Auto	1,479,266	594,409
5	acenaphthene d-10 (SURR)	12:25	745	Auto	1,323,814	845,312
6	phenanthrene d-10 (SURR)	17:07	1027	Auto	2,030,899	757,975
7	chrysene d-12 (SURR)	22:49	1369	Auto	1,479,266	594,409
8	1,3-dimethyl-2-nitrobenzene	8:05	485	Auto	836,385	477,719
9	triphenyl phosphate (S)	22:13	1333	Auto	1,153,682	460,281
10	perylene d-12 (SURR)	27:17	1637	Auto	834,377	147,651
11	hexachlorocyclopentadiene	10:16	616	Auto	2,115	1,255
12	bis(2-ethylhexyl)adipate	22:13	1333	Auto	29,160	5,612
13	bis(2-ethylhexyl)phthalate	23:07	1387	Auto	116,171	28,357
14	2,6-dinitrotoluene	12:04	724	Auto	34,852	19,793
15	2,4-dinitrotoluene	13:12	792	Auto	3,474	1,441

reviewed I.S./surrogate
 I.S. p-terphenyl d-14 Not Present.

Quantitation Report

Quanfile: 24472

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 10/30/01; SAMPLE 24472

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	745	12:25	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1027	17:07	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1369	22:49	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1369	22:49	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	745	12:25	BV	7.081	UG/L
6	phenanthrene d-10(SURR	A	1027	17:07	BV	7.407	UG/L
7	chrysene d-12 (SURR)	A	1369	22:49	BB	7.395	UG/L
8	1,3-dimethyl-2-nitrobe	A	485	8:05	BV	5.167	UG/L
9	triphenyl phosphate (S	A	1333	22:13	BB	5.722	UG/L
10	perylene d-12 (SURR)	A	1637	27:17	BB	4.486	UG/L
11	hexachlorocyclopentadi	A	616	10:16	BB	0.039	UG/L
16	bis(2-ethylhexyl)adipa	A	1333	22:13	MM	0.108	UG/L
17	bis(2-ethylhexyl)phtha	A	1387	23:07	VB	0.251	UG/L
20	2,6-dinitrotoluene	A	724	12:04	BB	0.418	UG/L
21	2,4-dinitrotoluene	A	792	13:12	BB	0.039	UG/L

Comments for Sample AD24472 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-21-2001 Time: 16:13:32

The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range, probably due to laboratory contamination.