


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/14/2001 Time: 10:52:02

Sample: AD24431
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/16/99 10:44 Ended: 10/31/01 10:44 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.2		1.0
20	Surr_Triphenyl phosphate		5.1		1.0
21	Surr_Perylene-d12		4.6		1.0

Chromatogram Plot File: E:\24431 Date: Oct-31-1991 20:53:22
 Comment: BOHLK; METHOD 525; INST 204; 10/31/01; SAMPLE 24431
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 5696844



Integration Report Quanfile: 24431 Quan Entries: 13
 Comment: BOHLK: METHOD 525; INST 204; 10/31/01; SAMPLE 24431
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:26	746	Auto	726,206	324,901
2	phenanthrene d-10 (IS)	17:07	1027	Auto	1,259,991	464,961
3	chrysene d-12 (IS)	22:50	1370	Auto	983,759	385,737
4	p-terphenyl d-14 (IS)	20:57	1257	Auto	3,131,460	1,752,270
5	acenaphthene d-10 (SURR)	12:26	746	Auto	726,206	324,901
6	phenanthrene d-10 (SURR)	17:07	1027	Auto	1,259,991	464,961
7	chrysene d-12 (SURR)	22:50	1370	Auto	983,759	385,737
8	1,3-dimethyl-2-nitrobenzene	8:06	486	Auto	371,468	151,736
9	triphenyl phosphate (S)	22:13	1333	Auto	684,535	211,413
10	perylene d-12 (SURR)	27:18	1638	Auto	564,150	93,735
11	bis(2-ethylhexyl)phthalate	23:08	1388	Auto	337,712	74,993
12	2,6-dinitrotoluene	12:05	725	Auto	26,009	15,688
13	4,4'-dichlorodiphenyl ether	20:52	1252	Auto	1,508	1,508

reversed IS/SURR's

Quantitation Report

Quanfile: 24431

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 10/31/01; SAMPLE 24431

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	746	12:26	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1027	17:07	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1370	22:50	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1257	20:57	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	746	12:26	BV	1.835	UG/L
6	phenanthrene d-10(SURR	A	1027	17:07	BV	2.171	UG/L
7	chrysene d-12 (SURR)	A	1370	22:50	BB	2.324	UG/L
8	1,3-dimethyl-2-nitrobe	A	486	8:06	BB	4.184	UG/L
9	triphenyl phosphate (S	A	1333	22:13	BB	5.106	UG/L
10	perylene d-12 (SURR)	A	1638	27:18	BB	4.561	UG/L
17	bis(2-ethylhexyl)phtha	A	1388	23:08	BV	1.097	UG/L
20	2,6-dinitrotoluene	A	725	12:05	BB	0.564	UG/L
28	4,4'-dde	A	1252	20:52	BB	0.014	UG/L

Comments for Sample AD24431 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-18-2001 Time: 09:59:24

The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range. Recoveries for 4 of the 18 compounds in one of the Spiked Samples were below their acceptance ranges.