

User: Bohk, Ted
Westchester Dept. of Labs & Research
Date: 11/17/2001 Time: 14:30:47

Sample: AD24477
Analysis: \$LP525 Organic Extractables
Units: ug
Started: 11/3/01 14:29 Ended: 11/3/01 14:29 Analyst: TB
Page: 1

	Component Name	Viol	Result
1	DFTPP		PASSED
2	ENDRIN <i>X</i>		30
3	DDT		2.1

flag.

Chromatogram Plot

File: E:\DFTA1103

Date: Nov-03-1991 13:30:22

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1859

Range: 1 to 1859

100% = 5755217



Chromatogram Plot

File: E:\DFTA1103 Date: Nov-03-1991 13:30:22


Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1080 Retention Time: 18:00 RIC: 37036 Mass Range: 47 - 86

Plotted: 1080 to 1120

Range: 1 to 1859

100% = 5755217



Mass List

File: E:\DFTA1103 Date: Nov-03-1991 13:30:22

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1104 Retention Time: 18:24 RIC: 5755217 Mass Range: 45 - 444

Peaks: 174 Mass of Base Peak: 442 Int: 968529 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
45	1,764	0.18	84	4,117	0.43	127	177,058	18.28
47	2,941	0.30	85	3,823	0.39	128	20,000	2.06
48	2,647	0.27	86	12,647	1.31	129	80,882	8.35
49	20,588	2.13	87	4,705	0.49	130	7,941	0.82
50	115,882	11.96	91	3,823	0.39	131	2,058	0.21
51	348,235	35.96	92	2,941	0.30	134	3,823	0.39
52	20,000	2.06	93	33,529	3.46	135	6,470	0.67
55	4,705	0.49	98	37,058	3.83	136	4,117	0.43
56	11,176	1.15	99	17,058	1.76	137	4,411	0.46
57	16,470	1.70	101	8,823	0.91	141	14,117	1.46
61	3,823	0.39	103	7,647	0.79	142	3,529	0.36
62	5,294	0.55	104	6,764	0.70	143	2,352	0.24
63	7,941	0.82	105	9,411	0.97	146	4,117	0.43
65	3,823	0.39	107	88,235	9.11	147	8,235	0.85
68	3,823	0.39	108	10,588	1.09	148	14,117	1.46
69	258,235	26.66	109	3,823	0.39	149	3,235	0.33
73	5,588	0.58	110	146,470	15.12	150	2,352	0.24
74	41,470	4.28	111	17,647	1.82	153	3,529	0.36
75	35,882	3.70	116	4,411	0.46	154	2,941	0.30
76	17,058	1.76	117	95,294	9.84	155	6,470	0.67
77	274,411	28.33	118	5,000	0.52	156	10,000	1.03
78	17,058	1.76	122	7,647	0.79	158	2,352	0.24
79	29,117	3.01	123	5,882	0.61	160	3,529	0.36
80	20,000	2.06	124	3,529	0.36	161	4,117	0.43
81	16,764	1.73	125	3,235	0.33	162	2,941	0.30

Mass List

File: E:\DFTA1103 Date: Nov-03-1991 13:30:22

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1104 Retention Time: 18:24 RIC: 5755217 Mass Range: 45 - 444

Peaks: 174 Mass of Base Peak: 442 Int: 968529 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
162	2,941	0.30	205	20,588	2.13	253	3,235	0.33
165	8,529	0.88	206	100,294	10.36	254	2,941	0.30
166	4,705	0.49	207	14,117	1.46	255	336,470	34.74
167	29,411	3.04	208	4,117	0.43	256	47,941	4.95
168	10,588	1.09	209	2,058	0.21	257	3,235	0.33
174	5,882	0.61	211	3,529	0.36	258	16,470	1.70
175	10,294	1.06	217	26,176	2.70	259	3,823	0.39
176	3,529	0.36	218	3,823	0.39	264	2,058	0.21
177	7,647	0.79	221	10,294	1.06	265	7,058	0.73
179	12,647	1.31	222	3,235	0.33	273	12,058	1.24
180	10,882	1.12	223	4,705	0.49	274	39,411	4.07
181	5,294	0.55	224	47,352	4.89	275	193,823	20.01
185	10,000	1.03	225	13,529	1.40	276	24,705	2.55
186	73,823	7.62	226	2,352	0.24	277	16,764	1.73
187	21,176	2.19	227	27,941	2.88	278	3,823	0.39
189	5,588	0.58	228	5,588	0.58	293	5,882	0.61
192	4,705	0.49	229	4,411	0.46	294	2,647	0.27
193	5,588	0.58	234	2,941	0.30	296	41,470	4.28
196	7,647	0.79	235	3,235	0.33	297	6,764	0.70
197	5,882	0.61	242	4,411	0.46	303	5,588	0.58
198	538,235	55.57	243	2,941	0.30	314	2,941	0.30
199	40,000	4.13	244	33,823	3.49	315	7,647	0.79
200	3,235	0.33	245	5,294	0.55	316	4,411	0.46
203	4,705	0.49	246	11,470	1.18	323	16,764	1.73
204	15,294	1.58	247	2,647	0.27	324	3,823	0.39

Mass List

File: E:\DFTA1103 Date: Nov-03-1991 13:30:22

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1104 Retention Time: 18:24 RIC: 5755217 Mass Range: 45 - 444

Peaks: 174 Mass of Base Peak: 442 Int: 968529 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
205	20,588	2.13	253	3,235	0.33	327	3,235	0.33
206	100,294	10.36	254	2,941	0.30	333	2,941	0.30
207	14,117	1.46	255	336,470	34.74	334	12,352	1.28
208	4,117	0.43	256	47,941	4.95	335	6,470	0.67
209	2,058	0.21	257	3,235	0.33	346	3,823	0.39
211	3,529	0.36	258	16,470	1.70	352	6,176	0.64
217	26,176	2.70	259	3,823	0.39	353	6,470	0.67
218	3,823	0.39	264	2,058	0.21	354	13,235	1.37
221	10,294	1.06	265	7,058	0.73	355	2,647	0.27
222	3,235	0.33	273	12,058	1.24	365	38,235	3.95
223	4,705	0.49	274	39,411	4.07	366	8,235	0.85
224	47,352	4.89	275	193,823	20.01	371	3,235	0.33
225	13,529	1.40	276	24,705	2.55	372	9,117	0.94
226	2,352	0.24	277	16,764	1.73	373	3,823	0.39
227	27,941	2.88	278	3,823	0.39	390	2,058	0.21
228	5,588	0.58	293	5,882	0.61	403	8,529	0.88
229	4,411	0.46	294	2,647	0.27	404	2,647	0.27
234	2,941	0.30	296	41,470	4.28	421	5,000	0.52
235	3,235	0.33	297	6,764	0.70	422	4,411	0.46
242	4,411	0.46	303	5,588	0.58	423	53,823	5.56
243	2,941	0.30	314	2,941	0.30	424	10,882	1.12
244	33,823	3.49	315	7,647	0.79	441	141,176	14.58
245	5,294	0.55	316	4,411	0.46	442	968,529	100.00
246	11,470	1.18	323	16,764	1.73	443	181,176	18.71
247	2,647	0.27	324	3,823	0.39	444	15,588	1.61

Chromatogram Plot

File: E:\DFTA1103

Date: Nov-03-1991 13:30:22

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1280

Retention Time: 21:20


RIC: 38407

Mass Range: 45 - 281

Plotted: 1260 to 1280

Range: 1 to 1859

100% = 5755217




Chromatogram Plot

File: E:\DFTA1103 Date: Nov-03-1991 13:30:22


Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN

Scan No: 1300 Retention Time: 21:40 RIC: 54218 Mass Range: 45 - 429

Plotted: 1285 to 1300 Range: 1 to 1859 100% = 5755217



Chromatogram Plot File: E:\DFTA1103 Date: Nov-03-1991 13:38:22
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DFTPP/ENDRIN
Scan No: 1370 Retention Time: 22:50 RIC: 34773 Mass Range: 45 - 281
Plotted: 1350 to 1370 Range: 1 to 1859 100% = 5755217



Chromatogram Plot

File: E:\DDA1103

Date: Nov-03-1991 14:29:33

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0


Plotted: 1 to 1860

Range: 1 to 1860

100% = 5939962



Chromatogram Plot File: E:\DDA1103 Date: Nov-03-1991 14:29:33
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DDT
 Scan No: 1257 Retention Time: 20:57 RIC: 33144 Mass Range: 47 - 207
 Plotted: 1237 to 1257 Range: 1 to 1860 100% = 5939962



Chromatogram Plot

File: E:\DDA1103

Date: Nov-03-1991 14:29:33

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DDT

Scan No: 1300

Retention Time: 21:40

RIC: 60483


Mass Range: 45 - 430

Plotted: 1270 to 1300

Range: 1 to 1860

100% = 5939962

2.09%



Chromatogram Plot File: E:\DDA1103 Date: Nov-03-1991 14:29:33
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; DDT
 Scan No: 1333 Retention Time: 22:13 RIC: 36180 Mass Range: 45 - 281
 Plotted: 1300 to 1333 Range: 1 to 1860 100% = 5939962

