


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/19/2001 Time: 10:19:29

Sample: AD24484
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/22/01 10:17 Ended: 11/3/01 10:17 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acefenchlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzene		5.3		1.0
20	Surr_Triphenyl phosphate		6.1		1.0
21	Surr_Perylene-d12		3.8		1.0

Chromatogram Plot File: C:\24484J Date: Nov-04-1991 13:55:32
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24484
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7015902



Integration Report Quanfile: 24484J Quan Entries: 12
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24484
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:20	740	Auto	1,611,515	1,031,318
2	phenanthrene d-10 (IS)	16:59	1019	Auto	2,590,055	998,153
3	chrysene d-12 (IS)	22:45	1365	Auto	1,872,054	680,547
4	p-terphenyl d-14 (IS)	20:53	1253	Auto	2,967,661	1,989,235
5	acenaphthene d-10 (SURR)	12:20	740	Auto	1,611,515	1,031,318
6	phenanthrene d-10 (SURR)	16:59	1019	Auto	2,590,055	998,153
7	chrysene d-12 (SURR)	22:45	1365	Auto	1,872,054	680,547
8	1,3-dimethyl-2-nitrobenzene	8:00	480	Auto	1,050,474	443,895
9	triphenyl phosphate (S)	22:09	1329	Auto	1,505,030	648,678
10	perylene d-12 (SURR)	27:07	1627	Auto	956,834	171,981
11	hexachlorocyclopentadiene	10:11	611	Auto	1,317	1,317
12	bis(2-ethylhexyl)phthalate	23:02	1382	Auto	696,030	199,184

reversed IS/Surr's

Quantitation Report

Quanfile: 24484J

Quan Entries: 12

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24484

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	740	12:20	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1019	16:59	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1253	20:53	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	740	12:20	BB	3.642	UG/L
6	phenanthrene d-10(SURR)	A	1019	16:59	BV	3.893	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	3.908	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	480	8:00	BV	5.252	UG/L
9	triphenyl phosphate (S)	A	1329	22:09	BB	6.104	UG/L
10	perylene d-12 (SURR)	A	1627	27:07	BB	3.767	UG/L
11	hexachlorocyclopentadiene	A	611	10:11	BB	0.018	UG/L
17	bis(2-ethylhexyl)phthalate	A	1382	23:02	VB	1.116	UG/L

Comments for Sample AD24484 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 13:12:12

The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 1.3 ug/L. Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.