


User: Bohik, Ted
Westchester Dept. of Labs & Research
Date: 11/19/2001 Time: 10:23:26

Sample: AD24485
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/22/01 10:20 Ended: 11/3/01 10:20 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.6		1.0
21	Surr_Perylene-d12		3.9		1.0

Chromatogram Plot File: C:\24485K Date: Nov-04-1991 14:29:57
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24485
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7126195



Integration Report Quanfile: 24485K Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24485
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:21	741	Auto	1,432,571	943,627
2	phenanthrene d-10 (IS)	17:00	1020	Auto	2,233,906	909,184
3	chrysene d-12 (IS)	22:46	1366	Auto	1,752,293	671,299
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,196,094	2,123,605
5	acenaphthene d-10 (SURR)	12:21	741	Auto	1,432,571	943,627
6	phenanthrene d-10 (SURR)	17:00	1020	Auto	2,233,906	909,184
7	chrysene d-12 (SURR)	22:46	1366	Auto	1,752,293	671,299
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	899,557	398,717
9	triphenyl phosphate (S)	22:10	1330	Auto	1,287,674	506,683
10	perylene d-12 (SURR)	27:08	1628	Auto	926,442	163,192
11	hexachlorobenzene	15:26	926	Auto	2,633	1,422
12	bis(2-ethylhexyl)adipate	22:10	1330	Auto	15,971	2,031
13	bis(2-ethylhexyl)phthalate	23:03	1383	Auto	256,319	56,645
14	4,4'-dichlorodiphenyl ether	20:49	1249	Auto	3,767	2,212

reversed IS/SURR's

Quantitation Report

Quanfile: 24485K

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24485

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	741	12:21	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1020	17:00	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1366	22:46	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	741	12:21	BB	3.006	UG/L
6	phenanthrene d-10(SURR	A	1020	17:00	BV	3.118	UG/L
7	chrysene d-12 (SURR)	A	1366	22:46	BB	3.397	UG/L
8	1,3-dimethyl-2-nitrobe	A	481	8:01	BV	5.059	UG/L
9	triphenyl phosphate (S	A	1330	22:10	BB	5.579	UG/L
10	perylene d-12 (SURR)	A	1628	27:08	BB	3.896	UG/L
12	hexachlorobenzene	A	926	15:26	BB	0.014	UG/L
16	bis(2-ethylhexyl)adipa	A	1330	22:10	MM	0.085	UG/L
17	bis(2-ethylhexyl)phtha	A	1383	23:03	VB	0.439	UG/L
28	4,4'-dde	A	1249	20:49	BB	0.016	UG/L

Comments for Sample AD24485 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 13:15:46

The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 1.3 ug/L. Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.