


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/17/2001 Time: 15:15:21

Sample: AD24488
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/20/01 15:12 Ended: 11/3/01 15:12 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.69		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.0		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21	Surr_Perylene-d12		3.8		1.0

Chromatogram Plot File: E:\24488C Date: Nov-04-1991 02:27:15
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24488
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 6811616



Integration Report Quanfile: 24488C Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24488
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:20	740	Auto	1,826,553	1,199,382
2	phenanthrene d-10 (IS)	16:59	1019	Auto	2,814,343	1,252,654
3	chrysene d-12 (IS)	22:44	1364	Auto	2,098,408	842,440
4	p-terphenyl d-14 (IS)	20:53	1253	Auto	3,461,179	2,055,508
5	acenaphthene d-10 (SURR)	12:20	740	Auto	1,826,553	1,199,382
6	phenanthrene d-10 (SURR)	16:59	1019	Auto	2,814,343	1,252,654
7	chrysene d-12 (SURR)	22:44	1364	Auto	2,098,408	842,440
8	1,3-dimethyl-2-nitrobenzene	8:00	480	Auto	1,141,103	596,454
9	triphenyl phosphate (S)	22:09	1329	Auto	1,515,236	527,514
10	perylene d-12 (SURR)	27:06	1626	Auto	1,089,278	198,848
11	hexachlorobenzene	15:24	924	Auto	7,450	3,058
12	bis(2-ethylhexyl)phthalate	23:02	1382	Auto	480,328	129,819
13	2,4-dinitrotoluene	13:07	787	Auto	9,297	3,295
14	butachlor	20:27	1227	Auto	0	0
15	4,4'-dichlorobiphenyl	20:48	1248	Auto	2,631	1,661

reversed IS/surrs

Quantitation Report Quanfile: 24488C Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24488
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	740	12:20	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1019	16:59	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1364	22:44	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1253	20:53	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	740	12:20	BB	3.539	UG/L
6	phenanthrene d-10 (SURR)	A	1019	16:59	BV	3.627	UG/L
7	chrysene d-12 (SURR)	A	1364	22:44	BB	3.756	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	480	8:00	BV	5.033	UG/L
9	triphenyl phosphate (S)	A	1329	22:09	BB	5.483	UG/L
10	perylene d-12 (SURR)	A	1626	27:06	BB	3.826	UG/L
12	hexachlorobenzene	A	924	15:24	BB	0.030	UG/L
17	bis(2-ethylhexyl)phthalate	A	1382	23:02	VB	0.686	UG/L
21	2,4-dinitrotoluene	A	787	13:07	BB	0.070	UG/L
27	butachlor	A	1227	20:27	BB	0.001	UG/L
28	4,4'-dichlorodiphenyl ether	A	1248	20:48	BB	0.009	UG/L

Comments for Sample AD24488 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 12:35:56

Recoveries for 2 of the 18 compounds in each of the MDL and LCS Standards were outside of their acceptance ranges.