


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/19/2001 Time: 10:29:11

Sample: AD24641
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/22/01 10:26 Ended: 11/3/01 10:26 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		6.2		1.0
21	Surr_Perylene-d12		4.0		1.0

Chromatogram Plot File: F:\246410 Date: Nov-04-1991 20:14:22
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24641
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 6791459



Integration Report Quanfile: 246410 Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24641
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:22	742	Auto	1,584,902	852,546
2	phenanthrene d-10 (IS)	17:01	1021	Auto	2,400,219	1,025,851
3	chrysene d-12 (IS)	22:46	1366	Auto	1,782,374	751,019
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,095,414	1,992,968
5	acenaphthene d-10 (SURR)	12:22	742	Auto	1,584,902	852,546
6	phenanthrene d-10 (SURR)	17:01	1021	Auto	2,400,219	1,025,851
7	chrysene d-12 (SURR)	22:46	1366	Auto	1,782,374	751,019
8	1,3-dimethyl-2-nitrobenzene	8:02	482	Auto	995,317	545,138
9	triphenyl phosphate (S)	22:10	1330	Auto	1,457,989	597,010
10	perylene d-12 (SURR)	27:09	1629	Auto	965,778	172,403
11	bis(2-ethylhexyl)adipate	22:07	1327	Auto	26,873	5,206
12	bis(2-ethylhexyl)phthalate	23:04	1384	Auto	207,296	48,885
13	2,6-dinitrotoluene	12:01	721	Auto	56,027	34,518

reversed IS/surr's

Quantitation Report Quanfile: 246410 Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24641
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	742	12:22	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1021	17:01	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1366	22:46	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	742	12:22	BB	3.434	UG/L
6	phenanthrene d-10(SURR)	A	1021	17:01	BV	3.459	UG/L
7	chrysene d-12 (SURR)	A	1366	22:46	BB	3.568	UG/L
8	1,3-dimethyl-2-nitrobe	A	482	8:02	BV	5.059	UG/L
9	triphenyl phosphate (S	A	1330	22:10	BB	6.211	UG/L
10	perylene d-12 (SURR)	A	1629	27:09	BB	3.993	UG/L
16	bis(2-ethylhexyl)adipa	A	1327	22:07	MM	0.140	UG/L
17	bis(2-ethylhexyl)phtha	A	1384	23:04	BB	0.349	UG/L
20	2,6-dinitrotoluene	A	721	12:01	BB	0.543	UG/L

Comments for Sample AD24641 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 13:46:43

The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 1.3 ug/L. Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.