


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/19/2001 Time: 10:34:28

Sample: AD24696
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/22/01 10:33 Ended: 11/3/01 10:33 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.2		1.0
20	Surr_Triphenyl phosphate		5.9		1.0
21	Surr_Perylene-d12		3.7		1.0

Chromatogram Plot File: F:\24696S Date: Nov-04-1991 22:32:05
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24696
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7234671



Integration Report Quanfile: 24696S Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24696
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:20	740	Auto	1,754,573	842,741
2	phenanthrene d-10 (IS)	17:00	1020	Auto	2,827,702	1,198,992
3	chrysene d-12 (IS)	22:45	1365	Auto	2,098,956	948,710
4	p-terphenyl d-14 (IS)	20:53	1253	Auto	3,188,696	2,263,877
5	acenaphthene d-10 (SURR)	12:20	740	Auto	1,754,573	842,741
6	phenanthrene d-10 (SURR)	17:00	1020	Auto	2,827,702	1,198,992
7	chrysene d-12 (SURR)	22:45	1365	Auto	2,098,956	948,710
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	1,123,175	583,540
9	triphenyl phosphate (S)	22:09	1329	Auto	1,641,840	761,083
10	perylene d-12 (SURR)	27:07	1627	Auto	1,058,715	188,417
11	hexachlorocyclopentadiene	10:11	611	Auto	1,509	1,509
12	hexachlorobenzene	15:25	925	Auto	11,434	4,798
13	bis(2-ethylhexyl)adipate	22:02	1322	Auto	35,129	5,675
14	bis(2-ethylhexyl)phthalate	23:03	1383	Auto	754,349	185,639
15	4,4'-dichlorodiphenyl ether	20:49	1249	Auto	8,535	4,838

reversed IS/ SURR's

Quantitation Report

Quanfile: 24696S

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24696

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	740	12:20	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1020	17:00	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1253	20:53	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	740	12:20	BB	3.690	UG/L
6	phenanthrene d-10(SURR	A	1020	17:00	BV	3.955	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	4.078	UG/L
8	1,3-dimethyl-2-nitrobe	A	481	8:01	BV	5.157	UG/L
9	triphenyl phosphate (S	A	1329	22:09	BB	5.939	UG/L
10	perylene d-12 (SURR)	A	1627	27:07	BB	3.717	UG/L
11	hexachlorocyclopentadi	A	611	10:11	BB	0.018	UG/L
12	hexachlorobenzene	A	925	15:25	BB	0.048	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.156	UG/L
17	bis(2-ethylhexyl)phtha	A	1383	23:03	BB	1.079	UG/L
28	4,4'-dde	A	1249	20:49	BB	0.029	UG/L

Comments for Sample AD24696 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 14:14:06

The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 1.3 ug/L. Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.