


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 11/19/2001 Time: 10:41:45

Sample: AD24730
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 10/22/01 10:38 Ended: 11/3/01 10:38 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.8		1.0
21	Surr_Perylene-d12		4.1		1.0

Chromatogram Plot File: F:\24730X Date: Nov-05-1991 01:24:14
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24730
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7203610



Integration Report Quanfile: 24730X Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24730
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:21	741	Auto	1,667,734	853,605
2	phenanthrene d-10 (IS)	17:00	1020	Auto	2,661,998	1,121,462
3	chrysene d-12 (IS)	22:45	1365	Auto	1,937,165	795,554
4	p-terphenyl d-14 (IS)	20:53	1253	Auto	3,429,930	2,289,679
5	acenaphthene d-10 (SURR)	12:21	741	Auto	1,667,734	853,605
6	phenanthrene d-10 (SURR)	17:00	1020	Auto	2,661,998	1,121,462
7	chrysene d-12 (SURR)	22:45	1365	Auto	1,937,165	795,554
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	1,056,462	569,206
9	triphenyl phosphate (S)	22:09	1329	Auto	1,478,480	610,836
10	perylene d-12 (SURR)	27:08	1628	Auto	1,078,008	212,608
11	bis(2-ethylhexyl)adipate	22:02	1322	Auto	31,738	5,235
12	bis(2-ethylhexyl)phthalate	23:03	1383	Auto	362,892	99,420
13	2,4-dinitrotoluene	13:08	788	Auto	2,256	1,258

reversed IS/ SURR's

Quantitation Report

Quanfile: 24730X

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24730

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	741	12:21	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1020	17:00	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1253	20:53	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	741	12:21	BV	3.261	UG/L
6	phenanthrene d-10(SURR)	A	1020	17:00	BV	3.462	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	3.499	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	481	8:01	BV	5.104	UG/L
9	triphenyl phosphate (S)	A	1329	22:09	BB	5.795	UG/L
10	perylene d-12 (SURR)	A	1628	27:08	BB	4.101	UG/L
16	bis(2-ethylhexyl)adipate	A	1322	22:02	MM	0.152	UG/L
17	bis(2-ethylhexyl)phthalate	A	1383	23:03	BB	0.562	UG/L
21	2,4-dinitrotoluene	A	788	13:08	BB	0.019	UG/L

Comments for Sample AD24730 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-26-2001 Time: 10:51:49

The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 1.3 ug/L. Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.