


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/19/2001 Time: 10:43:04

Sample: AD24731
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/22/01 10:41 Ended: 11/3/01 10:41 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.0		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.2		1.0
20	Surr_Triphenyl phosphate		5.9		1.0
21	Surr_Perylene-d12		3.7		1.0

Chromatogram Plot File: F:\24731Y Date: Nov-05-1991 01:58:40
Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24731
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7661357



Integration Report Quanfile: 24731Y Quan Entries: 12
 Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24731
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Node	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:22	742	Auto	1,397,823	750,206
2	phenanthrene d-10 (IS)	17:01	1021	Auto	2,269,532	984,422
3	chrysene d-12 (IS)	22:46	1366	Auto	1,793,041	761,211
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,111,415	2,170,227
5	acenaphthene d-10 (SURR)	12:22	742	Auto	1,397,823	750,206
6	phenanthrene d-10 (SURR)	17:01	1021	Auto	2,269,532	984,422
7	chrysene d-12 (SURR)	22:46	1366	Auto	1,793,041	761,211
8	1,3-dimethyl-2-nitrobenzene	8:02	482	Auto	908,726	467,282
9	triphenyl phosphate (S)	22:10	1330	Auto	1,385,682	625,489
10	perylene d-12 (SURR)	27:09	1629	Auto	911,393	168,871
11	bis(2-ethylhexyl)adipate	22:03	1323	Auto	73,255	21,349
12	bis(2-ethylhexyl)phthalate	23:04	1384	Auto	1,174,006	364,389

reversed IS/SURR's

Quantitation Report

Quanfile: 24731Y

Quan Entries: 12

Comment: BOHLK; METHOD 525; INST 204; 11/03/01; SAMPLE 24731

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	742	12:22	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1021	17:01	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1366	22:46	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	742	12:22	BB	3.013	UG/L
6	phenanthrene d-10 (SURR)	A	1021	17:01	BV	3.254	UG/L
7	chrysene d-12 (SURR)	A	1366	22:46	BB	3.570	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	482	8:02	BV	5.237	UG/L
9	triphenyl phosphate (S)	A	1330	22:10	BV	5.868	UG/L
10	perylene d-12 (SURR)	A	1629	27:09	BB	3.746	UG/L
16	bis(2-ethylhexyl)adipate	A	1323	22:03	MM	0.379	UG/L
17	bis(2-ethylhexyl)phthalate	A	1384	23:04	BB	1.975	UG/L

Comments for Sample AD24731 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-26-2001 Time: 10:55:30

The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 1.3 ug/L. Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.