


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/21/2001 Time: 12:51:34


Sample: AD24982
Analysis: \$LP525 Organic Extractables
Units: ug
Started: 11/7/01 12:49 Ended: 11/7/01 12:49 Analyst: TB
Page: 1

	Component Name	Viol	Result
1	DFTPP		PASSED
2	ENDRIN	USPEC	22
3	DDT		3.1

Chromatogram Plot File: E:\DFTA1107 Date: Nov-07-1991 16:30:06
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5962262



Chromatogram Plot File: E:\DFTA1107 Date: Nov-07-1991 16:30:06
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1080 Retention Time: 18:00 RIC: 30039 Mass Range: 47 - 86
Plotted: 1080 to 1120 Range: 1 to 1860 100% = 5962262



Mass List

File: E:\DFTA1107 Date: Nov-07-1991 16:30:06

Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN

Scan No: 1104 Retention Time: 18:24 RIC: 5962262 Mass Range: 47 - 444

Peaks: 166 Mass of Base Peak: 442 Int: 1037987 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	2,922	0.28	84	4,220	0.41	124	4,870	0.47
49	15,909	1.53	85	3,571	0.34	127	192,857	18.58
50	118,831	11.45	86	10,389	1.00	128	14,935	1.44
51	373,376	35.97	88	3,246	0.31	129	82,142	7.91
52	18,506	1.78	91	4,545	0.44	130	6,168	0.59
55	5,844	0.56	92	4,545	0.44	135	6,493	0.63
56	10,389	1.00	93	35,714	3.44	136	3,571	0.34
57	18,831	1.81	94	3,571	0.34	137	4,220	0.41
61	7,142	0.69	98	47,727	4.60	141	12,337	1.19
62	4,870	0.47	99	19,480	1.88	142	4,545	0.44
63	11,688	1.13	100	3,246	0.31	143	2,922	0.28
65	3,246	0.31	101	6,493	0.63	147	8,441	0.81
68	5,194	0.50	103	7,142	0.69	148	17,207	1.66
69	272,077	26.21	104	7,467	0.72	149	4,220	0.41
73	7,142	0.69	105	7,467	0.72	153	4,545	0.44
74	48,376	4.66	106	2,597	0.25	154	4,545	0.44
75	42,207	4.07	107	87,662	8.45	155	8,116	0.78
76	11,363	1.09	108	8,766	0.84	156	7,792	0.75
77	265,584	25.59	110	144,805	13.95	160	4,220	0.41
78	16,233	1.56	111	20,779	2.00	161	8,766	0.84
79	29,545	2.85	116	4,870	0.47	162	1,948	0.19
80	17,857	1.72	117	112,662	10.85	163	1,623	0.16
81	18,181	1.75	118	6,818	0.66	165	5,519	0.53
82	3,571	0.34	122	5,519	0.53	166	5,194	0.50
83	4,220	0.41	123	7,142	0.69	167	30,844	2.97

Mass List

File: E:\DFTA1107 Date: Nov-07-1991 16:30:06

Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN

Scan No: 1104 Retention Time: 18:24 RIC: 5962262 Mass Range: 47 - 444

Peaks: 166 Mass of Base Peak: 442 Int: 1037987 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
166	5,194	0.50	210	3,246	0.31	258	18,831	1.81
167	30,844	2.97	211	5,519	0.53	259	4,220	0.41
168	10,714	1.03	215	3,246	0.31	265	7,142	0.69
174	4,545	0.44	217	28,896	2.78	273	14,935	1.44
175	9,415	0.91	218	3,896	0.38	274	45,779	4.41
177	6,168	0.59	221	11,688	1.13	275	176,623	17.02
179	12,987	1.25	222	3,896	0.38	276	24,025	2.31
180	7,467	0.72	223	5,519	0.53	277	15,909	1.53
181	3,246	0.31	224	53,571	5.16	282	1,948	0.19
184	3,896	0.38	225	13,961	1.35	293	6,818	0.66
185	9,740	0.94	226	2,597	0.25	296	41,558	4.00
186	71,753	6.91	227	26,948	2.60	297	6,493	0.63
187	19,805	1.91	228	3,571	0.34	303	7,142	0.69
189	3,896	0.38	229	4,220	0.41	304	3,571	0.34
192	4,870	0.47	231	4,545	0.44	314	3,246	0.31
193	5,519	0.53	236	2,922	0.28	315	7,467	0.72
196	7,467	0.72	243	3,571	0.34	316	4,220	0.41
198	578,246	55.71	244	36,038	3.47	323	21,753	2.10
199	35,714	3.44	245	5,194	0.50	327	3,246	0.31
203	4,220	0.41	246	12,662	1.22	328	2,922	0.28
204	15,909	1.53	250	2,597	0.25	334	14,610	1.41
205	23,051	2.22	253	1,623	0.16	335	5,844	0.56
206	99,350	9.57	255	363,961	35.06	346	5,844	0.56
207	10,389	1.00	256	48,376	4.66	352	6,818	0.66
208	4,545	0.44	257	4,870	0.47	353	5,519	0.53

197

Mass List

File: E:\DFTA1107 Date: Nov-07-1991 16:30:06


Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN

Scan No: 1104 Retention Time: 18:24 RIC: 5962262 Mass Range: 47 - 444


Peaks: 166 Mass of Base Peak: 442 Int: 1037987 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
199	35,714	3.44	245	5,194	0.50	327	3,246	0.31
203	4,220	0.41	246	12,662	1.22	328	2,922	0.28
204	15,909	1.53	250	2,597	0.25	334	14,610	1.41
205	23,051	2.22	253	1,623	0.16	335	5,844	0.56
206	99,350	9.57	255	363,961	35.06	346	5,844	0.56
207	10,389	1.00	256	48,376	4.66	352	6,818	0.66
208	4,545	0.44	257	4,870	0.47	353	5,519	0.53
210	3,246	0.31	258	18,831	1.81	354	9,415	0.91
211	5,519	0.53	259	4,220	0.41	355	3,246	0.31
215	3,246	0.31	265	7,142	0.69	365	47,402	4.57
217	28,896	2.78	273	14,935	1.44	366	5,519	0.53
218	3,896	0.38	274	45,779	4.41	372	13,311	1.28
221	11,688	1.13	275	176,623	17.02	373	4,870	0.47
222	3,896	0.38	276	24,025	2.31	390	2,272	0.22
223	5,519	0.53	277	15,909	1.53	402	4,220	0.41
224	53,571	5.16	282	1,948	0.19	403	6,493	0.63
225	13,961	1.35	293	6,818	0.66	404	4,220	0.41
226	2,597	0.25	296	41,558	4.00	421	5,844	0.56
227	26,948	2.60	297	6,493	0.63	422	8,766	0.84
228	3,571	0.34	303	7,142	0.69	423	54,545	5.25
229	4,220	0.41	304	3,571	0.34	424	12,987	1.25
231	4,545	0.44	314	3,246	0.31	441	133,116	12.82
236	2,922	0.28	315	7,467	0.72	442	1,037,987	100.00
243	3,571	0.34	316	4,220	0.41	443	178,246	17.17
244	36,038	3.47	323	21,753	2.10	444	11,688	1.13


Chromatogram Plot File: E:\DFTA1107 Date: Nov-07-1991 16:30:06
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
 Scan No: 1280 Retention Time: 21:20 RIC: 36004 Mass Range: 47 - 355
 Plotted: 1260 to 1280 Range: 1 to 1860 100% = 5962262




Chromatogram Plot File: E:\DFTA1107 Date: Nov-07-1991 16:30:06
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
 Scan No: 1302 Retention Time: 21:42 RIC: 45476 Mass Range: 45 - 429
 Plotted: 1282 to 1302 Range: 1 to 1860 100% = 5962262




Chromatogram Plot File: E:\DFTA1107 Date: Nov-07-1991 16:30:06
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 32563 Mass Range: 45 - 281
Plotted: 1347 to 1367 Range: 1 to 1860 100% = 5962262
8.68%




Chromatogram Plot File: E:\DDTA1107 Date: Nov-07-1991 17:04:28
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 4843747




Chromatogram Plot File: E:\DDTA1107 Date: Nov-07-1991 17:04:28
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
Scan No: 1277 Retention Time: 21:17 RIC: 32155 Mass Range: 47 - 207
Plotted: 1237 to 1277 Range: 1 to 1060 100% = 4843747




Chromatogram Plot File: E:\DDTA1107 Date: Nov-07-1991 17:04:28
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
Scan No: 1292 Retention Time: 21:32 RIC: 38933 Mass Range: 47 - 237
Plotted: 1272 to 1292 Range: 1 to 1860 100% = 4843747
2.52%




Chromatogram Plot File: E:\DDTA1107 Date: Nov-07-1991 17:04:28
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
Scan No: 1320 Retention Time: 22:00 RIC: 44555 Mass Range: 45 - 281
Plotted: 1305 to 1320 Range: 1 to 1860 100% = 4843747



Chromatogram Plot File: C:\DFTB1107 Date: Nov-08-1991 04:35:26
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 5313744



Chromatogram Plot File: C:\DFTB1107 Date: Nov-08-1991 04:35:26
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1080 Retention Time: 18:00 RIC: 51571 Mass Range: 47 - 86
Plotted: 1080 to 1120 Range: 1 to 1859 100% = 5313744



Mass List

File: C:\DFTB1107 Date: Nov-08-1991 04:35:26

Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN

Scan No: 1103 Retention Time: 18:23 RIC: 5313744 Mass Range: 47 - 444

Peaks: 173 Mass of Base Peak: 442 Int: 892819 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	2,925	0.33	83	5,053	0.57	124	4,521	0.51
48	3,723	0.42	84	6,648	0.74	125	3,457	0.39
49	23,936	2.68	85	3,191	0.36	127	163,297	18.29
50	125,531	14.06	86	11,968	1.34	128	14,361	1.61
51	335,372	37.56	87	3,457	0.39	129	68,617	7.69
52	17,021	1.91	91	5,585	0.63	130	7,180	0.80
55	3,191	0.36	92	4,787	0.54	134	2,127	0.24
56	9,308	1.04	93	34,308	3.84	135	7,180	0.80
57	16,489	1.85	94	2,393	0.27	136	2,925	0.33
61	7,446	0.83	98	33,244	3.72	137	3,723	0.42
62	7,180	0.80	99	14,361	1.61	141	10,106	1.13
63	8,776	0.98	101	5,585	0.63	142	4,787	0.54
65	2,925	0.33	103	5,319	0.60	143	2,659	0.30
68	5,053	0.57	104	5,319	0.60	146	3,723	0.42
69	235,106	26.33	105	7,978	0.89	147	5,585	0.63
73	5,851	0.66	106	2,393	0.27	148	19,680	2.20
74	35,106	3.93	107	72,074	8.07	149	3,191	0.36
75	36,436	4.08	108	10,106	1.13	153	3,989	0.45
76	15,425	1.73	110	140,425	15.73	154	3,723	0.42
77	220,478	24.69	111	14,361	1.61	155	7,978	0.89
78	14,627	1.64	116	3,191	0.36	156	8,776	0.98
79	32,180	3.60	117	96,276	10.78	157	2,659	0.30
80	16,223	1.82	118	7,712	0.86	158	2,659	0.30
81	11,702	1.31	122	6,382	0.71	159	2,659	0.30
82	3,723	0.42	123	6,648	0.74	160	4,255	0.48

Mass List

File: C:\DFTB1107 Date: Nov-08-1991 04:35:26

Comment: BOHLK: METHOD 525: INST 204: 11/07/01: DFTPP/ENDRIN

Scan No: 1103 Retention Time: 18:23 RIC: 5313744 Mass Range: 47 - 444

Peaks: 173 Mass of Base Peak: 442 Int: 892819 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
161	4,521	0.51	206	83,510	9.35	255	339,627	38.04
165	3,989	0.45	207	9,840	1.10	256	43,085	4.83
167	26,861	3.01	208	5,585	0.63	257	3,191	0.36
168	7,180	0.80	211	3,191	0.36	258	13,563	1.52
169	2,925	0.33	217	22,340	2.50	259	3,191	0.36
174	3,457	0.39	218	2,659	0.30	265	8,244	0.92
175	9,042	1.01	220	2,127	0.24	273	10,106	1.13
177	4,255	0.48	221	7,712	0.86	274	32,712	3.66
179	14,361	1.61	223	3,989	0.45	275	182,446	20.43
180	6,914	0.77	224	50,531	5.66	276	21,808	2.44
181	4,255	0.48	225	11,702	1.31	277	14,361	1.61
185	7,712	0.86	226	2,925	0.33	278	2,659	0.30
186	67,021	7.51	227	25,797	2.89	285	2,659	0.30
187	14,627	1.64	228	3,457	0.39	293	5,053	0.57
188	3,723	0.42	229	5,319	0.60	296	35,904	4.02
189	3,723	0.42	233	2,659	0.30	297	5,053	0.57
192	2,659	0.30	234	2,659	0.30	303	4,787	0.54
193	3,989	0.45	236	2,393	0.27	315	5,585	0.63
196	8,776	0.98	242	2,659	0.30	316	3,457	0.39
197	9,308	1.04	243	2,393	0.27	323	18,617	2.09
198	494,148	55.35	244	35,904	4.02	324	3,723	0.42
199	35,372	3.96	245	6,914	0.77	327	2,659	0.30
200	2,925	0.33	246	11,968	1.34	334	9,574	1.07
204	15,159	1.70	247	2,393	0.27	335	4,255	0.48
205	24,468	2.74	254	1,595	0.18	336	1,595	0.18

Mass List

File: C:\DFTB1107 Date: Nov-08-1991 04:35:26

Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN

Scan No: 1103 Retention Time: 18:23 RIC: 5313744 Mass Range: 47 - 444

Peaks: 173 Mass of Base Peak: 442 Int: 892819 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
161	4,521	0.51	206	83,510	9.35	255	339,627	38.04
165	3,989	0.45	207	9,840	1.10	256	43,085	4.83
167	26,861	3.01	208	5,585	0.63	257	3,191	0.36
168	7,180	0.80	211	3,191	0.36	258	13,563	1.52
169	2,925	0.33	217	22,340	2.50	259	3,191	0.36
174	3,457	0.39	218	2,659	0.30	265	8,244	0.92
175	9,042	1.01	220	2,127	0.24	273	10,106	1.13
177	4,255	0.48	221	7,712	0.86	274	32,712	3.66
179	14,361	1.61	223	3,989	0.45	275	182,446	20.43
180	6,914	0.77	224	50,531	5.66	276	21,808	2.44
181	4,255	0.48	225	11,702	1.31	277	14,361	1.61
185	7,712	0.86	226	2,925	0.33	278	2,659	0.30
186	67,021	7.51	227	25,797	2.89	285	2,659	0.30
187	14,627	1.64	228	3,457	0.39	293	5,053	0.57
188	3,723	0.42	229	5,319	0.60	296	35,904	4.02
189	3,723	0.42	233	2,659	0.30	297	5,053	0.57
192	2,659	0.30	234	2,659	0.30	303	4,787	0.54
193	3,989	0.45	236	2,393	0.27	315	5,585	0.63
196	8,776	0.98	242	2,659	0.30	316	3,457	0.39
197	9,308	1.04	243	2,393	0.27	323	18,617	2.09
198	494,148	55.35	244	35,904	4.02	324	3,723	0.42
199	35,372	3.96	245	6,914	0.77	327	2,659	0.30
200	2,925	0.33	246	11,968	1.34	334	9,574	1.07
204	15,159	1.70	247	2,393	0.27	335	4,255	0.48
205	24,468	2.74	254	1,595	0.18	336	1,595	0.18

Mass List

File: C:\DFTB1107 Date: Nov-08-1991 04:35:26


Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN

Scan No: 1103 Retention Time: 18:23 RIC: 5313744 Mass Range: 47 - 444


Peaks: 173 Mass of Base Peak: 442 Int: 892819 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
204	15,159	1.70	247	2,393	0.27	335	4,255	0.48
205	24,468	2.74	254	1,595	0.18	336	1,595	0.18
206	83,510	9.35	255	339,627	38.04	341	2,925	0.33
207	9,840	1.10	256	43,085	4.83	346	4,255	0.48
208	5,585	0.63	257	3,191	0.36	352	5,851	0.66
211	3,191	0.36	258	13,563	1.52	353	7,180	0.80
217	22,340	2.50	259	3,191	0.36	354	11,170	1.25
218	2,659	0.30	265	8,244	0.92	365	35,106	3.93
220	2,127	0.24	273	10,106	1.13	366	7,180	0.80
221	7,712	0.86	274	32,712	3.66	371	3,191	0.36
223	3,989	0.45	275	182,446	20.43	372	11,702	1.31
224	50,531	5.66	276	21,808	2.44	373	4,255	0.48
225	11,702	1.31	277	14,361	1.61	383	3,457	0.39
226	2,925	0.33	278	2,659	0.30	391	2,393	0.27
227	25,797	2.89	285	2,659	0.30	402	3,989	0.45
228	3,457	0.39	293	5,053	0.57	403	6,382	0.71
229	5,319	0.60	296	35,904	4.02	404	2,393	0.27
233	2,659	0.30	297	5,053	0.57	421	4,521	0.51
234	2,659	0.30	303	4,787	0.54	422	5,319	0.60
236	2,393	0.27	315	5,585	0.63	423	44,414	4.97
242	2,659	0.30	316	3,457	0.39	424	10,106	1.13
243	2,393	0.27	323	18,617	2.09	441	127,925	14.33
244	35,904	4.02	324	3,723	0.42	442	892,819	100.00
245	6,914	0.77	327	2,659	0.30	443	163,297	18.29
246	11,968	1.34	334	9,574	1.07	444	14,095	1.58


Chromatogram Plot File: C:\DFTB1107 Date: Nov-08-1991 04:35:26
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
 Scan No: 1280 Retention Time: 21:20 RIC: 52063 Mass Range: 47 - 355
 Plotted: 1260 to 1280 Range: 1 to 1859 100% = 5313744




Chromatogram Plot File: C:\DFTB1107 Date: Nov-08-1991 04:35:26
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1302 Retention Time: 21:42 RIC: 69261 Mass Range: 45 - 429
Plotted: 1282 to 1302 Range: 1 to 1859 100% = 5313744
8.97%




Chromatogram Plot File: C:\DFTB1107 Date: Nov-08-1991 04:35:26
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 50378 Mass Range: 45 - 355
Plotted: 1347 to 1367 Range: 1 to 1859 100% = 5313744




Chromatogram Plot File: F:\DDTB1107 Date: Nov-08-1991 05:09:49
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 5076612




Chromatogram Plot File: F:\DDTB1107 Date: Nov-08-1991 05:09:49
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
Scan No: 1259 Retention Time: 20:59 RIC: 42118 Mass Range: 47 - 207
Plotted: 1239 to 1259 Range: 1 to 1860 100% = 5076612
2.98%



Chromatogram Plot File: F:\DDTB1107 Date: Nov-08-1991 05:09:49
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
Scan No: 1292 Retention Time: 21:32 RIC: 53074 Mass Range: 45 - 237
Plotted: 1272 to 1292 Range: 1 to 1860 100% = 5076612
3.71%



Chromatogram Plot File: F:\DDTB1107 Date: Nov-08-1991 05:09:49
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; DDT
Scan No: 1324 Retention Time: 22:04 RIC: 49783 Mass Range: 47 - 235
Plotted: 1304 to 1324 Range: 1 to 1860 100% = 5076612



Breakdown = 4.3%