


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/21/2001 Time: 14:30:16

Sample: AD24732
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/23/01 14:28 Ended: 11/7/01 14:28 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.4		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.8		1.0
21	Surr_Perylene-d12		4.4		1.0

Chromatogram Plot File: E:\24732A Date: Nov-07-1991 20:33:49
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; SAMPLE 24732
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7660114



Integration Report Quanfile: 24732A Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; SAMPLE 24732
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:21	741	Auto	1,512,761	974,019
2	phenanthrene d-10 (IS)	17:00	1020	Auto	2,381,523	928,499
3	chrysene d-12(IS)	22:45	1365	Auto	1,794,265	624,071
4	p-terphenyl d-14(IS)	20:54	1254	Auto	3,694,505	2,329,629
5	acenaphthene d-10(SURR)	12:21	741	Auto	1,512,761	974,019
6	phenanthrene d-10(SURR)	17:00	1020	Auto	2,381,523	928,499
7	chrysene d-12 (SURR)	22:45	1365	Auto	1,794,265	624,071
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	958,456	504,166
9	triphenyl phosphate (S)	22:10	1330	Auto	1,362,259	539,144
10	perylene d-12 (SURR)	27:07	1627	Auto	1,070,202	203,644
11	bis(2-ethylhexyl)adipate	22:02	1322	Auto	22,783	2,969
12	bis(2-ethylhexyl)phthalate	23:03	1383	Auto	848,197	275,127
13	2,6-dinitrotoluene	12:01	721	Auto	57,748	30,845
14	2,4-dinitrotoluene	13:08	788	Auto	5,328	2,187
15	butachlor	20:28	1228	Auto	0	0
16	4,4'-dichlorodiphenyl ether	20:49	1249	Auto	2,829	1,516

reviewed IS/ Surrogate's

Quantitation Report

Quanfile: 24732A

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/07/01; SAMPLE 24732

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	741	12:21	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1020	17:00	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	741	12:21	BB	2.746	UG/L
6	phenanthrene d-10(SURR	A	1020	17:00	BV	2.875	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	3.009	UG/L
8	1,3-dimethyl-2-nitroben	A	481	8:01	BB	5.104	UG/L
9	triphenyl phosphate (S	A	1330	22:10	BB	5.764	UG/L
10	perylene d-12 (SURR)	A	1627	27:07	BB	4.396	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.118	UG/L
17	bis(2-ethylhexyl)phtha	A	1383	23:03	VB	1.421	UG/L
20	2,6-dinitrotoluene	A	721	12:01	BB	0.586	UG/L
21	2,4-dinitrotoluene	A	788	13:00	BB	0.049	UG/L
27	butachlor	A	1228	20:28	BB	0.001	UG/L
28	4,4'-dde	A	1249	20:49	BB	0.012	UG/L

Comments for Sample AD24732 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 09:41:15

Recoveries for 7 of the 18 compounds in the MDL Standard were outside of their acceptance ranges, with no recovery for Benzo(a)pyrene, probably due to contamination showing in its chromatographic region. Recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the Spiked Sample and Spiked Duplicate Sample were outside of their acceptance ranges.