


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/21/2001 Time: 15:35:35

Sample: AD24877
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/23/01 15:34 Ended: 11/7/01 15:34 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		6.1		1.0
21	Surr_Perylene-d12		3.6		1.0

Chromatogram Plot File: C:\24877H Date: Nov-08-1991 02:17:51
Comment: BOHLK; METHOD 525; INST 204; 11/07/01; SAMPLE 24877
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 8321500



Integration Report Quanfile: 24877H Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; SAMPLE 24877
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Node	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:20	740	Auto	1,713,926	1,081,451
2	phenanthrene d-10 (IS)	16:59	1019	Auto	2,792,266	1,176,150
3	chrysene d-12 (IS)	22:45	1365	Auto	1,950,326	734,175
4	p-terphenyl d-14 (IS)	20:53	1253	Auto	3,602,598	2,590,370
5	acenaphthene d-10 (SURR)	12:20	740	Auto	1,713,926	1,081,451
6	phenanthrene d-10 (SURR)	16:59	1019	Auto	2,792,266	1,176,150
7	chrysene d-12 (SURR)	22:45	1365	Auto	1,950,326	734,175
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	1,082,978	454,430
9	triphenyl phosphate (S)	22:09	1329	Auto	1,573,306	718,440
10	perylene d-12 (SURR)	27:06	1626	Auto	946,579	168,445
11	hexachlorocyclopentadiene	10:11	611	Auto	1,797	1,797
12	bis(2-ethylhexyl)adipate	22:01	1321	Auto	37,070	6,726
13	bis(2-ethylhexyl)phthalate	23:02	1382	Auto	225,615	62,463
14	4,4'-dichlorodiphenyl ether	20:49	1249	Auto	3,587	1,863

reversed IS/ SURR

Quantitation Report Quanfile: 24877H Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/07/01; SAMPLE 24877
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	740	12:20	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1019	16:59	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1253	20:53	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	740	12:20	BV	3.191	UG/L
6	phenanthrene d-10(SURR	A	1019	16:59	BV	3.457	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	3.354	UG/L
8	1,3-dimethyl-2-nitrobe	A	481	8:01	BV	5.091	UG/L
9	triphenyl phosphate (S	A	1329	22:09	BB	6.125	UG/L
10	perylene d-12 (SURR)	A	1626	27:06	BB	3.577	UG/L
11	hexachlorocyclopentadi	A	611	10:11	BB	0.022	UG/L
16	bis(2-ethylhexyl)adipa	A	1321	22:01	MM	0.177	UG/L
17	bis(2-ethylhexyl)phtha	A	1382	23:02	VB	0.347	UG/L
28	4,4'-dde	A	1249	20:49	BB	0.013	UG/L

Comments for Sample AD24877 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-24-2001 Time: 10:48:10

Recoveries for 7 of the 18 compounds in the MDL Standard were outside of their acceptance ranges, with no recovery for Benzo(a)pyrene, probably due to contamination showing in its chromatographic region. Recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the Spiked Sample and Spiked Duplicate Sample were outside of their acceptance ranges.