


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/24/2001 Time: 14:52:30

Sample: AD24948
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/24/01 14:50 Ended: 11/8/01 14:50 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.3		1.0
20	Surr_Triphenyl phosphate		5.7		1.0
21	Surr_Perylene-d12		3.1		1.0

Chromatogram Plot File: E:\24948A Date: Nov-08-1991 21:30:42
Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 24948
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 7386258



Integration Report Quanfile: 24948A Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 24948
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:21	741	Auto	1,709,875	979,040
2	phenanthrene d-10 (IS)	17:00	1020	Auto	2,884,964	1,257,009
3	chrysene d-12 (IS)	22:45	1365	Auto	1,817,536	735,612
4	p-terphenyl d-14 (IS)	20:53	1253	Auto	3,938,324	2,423,344
5	acenaphthene d-10 (SURR)	12:21	741	Auto	1,709,875	979,040
6	phenanthrene d-10 (SURR)	17:00	1020	Auto	2,884,964	1,257,009
7	chrysene d-12 (SURR)	22:45	1365	Auto	1,817,536	735,612
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	1,132,383	628,165
9	triphenyl phosphate (S)	22:09	1329	Auto	1,368,212	465,090
10	perylene d-12 (SURR)	27:08	1628	Auto	763,077	136,957
11	hexachlorocyclopentadiene	10:12	612	Auto	2,633	1,522
12	hexachlorobenzene	15:26	926	Auto	1,296	1,296
13	bis(2-ethylhexyl)adipate	22:02	1322	Auto	21,685	2,039
14	bis(2-ethylhexyl)phthalate	23:03	1383	Auto	338,365	86,095
15	4,4'-dichlorodiphenyl ether	20:49	1249	Auto	1,322	1,322

reviewed IS/surr's

Quantitation Report

Quanfile: 24948A

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 24948

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	741	12:21	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1020	17:00	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1253	20:53	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	741	12:21	BV	2.912	UG/L
6	phenanthrene d-10 (SURR)	A	1020	17:00	BV	3.267	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	2.859	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	481	8:01	BV	5.335	UG/L
9	triphenyl phosphate (S)	A	1329	22:09	BB	5.716	UG/L
10	perylene d-12 (SURR)	A	1628	27:08	BB	3.094	UG/L
11	hexachlorocyclopentadiene	A	612	10:12	BB	0.033	UG/L
12	hexachlorobenzene	A	926	15:26	BB	0.006	UG/L
16	bis(2-ethylhexyl)adipate	A	1322	22:02	MM	0.111	UG/L
17	bis(2-ethylhexyl)phthalate	A	1383	23:03	VB	0.558	UG/L
28	4,4'-dichlorodiphenyl ether	A	1249	20:49	BB	0.005	UG/L

Comments for Sample AD24948 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-28-2001 Time: 14:24:02

Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.