


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/26/2001 Time: 08:56:43

Sample: AD24949
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/24/01 14:52 Ended: 11/8/01 14:52 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.69		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.0		1.0
20	Surr_Triphenyl phosphate		5.8		1.0
21	Surr_Perylene-d12		3.2		1.0

Chromatogram Plot File: C:\24949B Date: Nov-08-1991 22:05:06
Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 24949
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 6141283



Integration Report

Quanfile: 24949B

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 24949

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:22	742	Auto	1,516,015	890,887
2	phenanthrene d-10 (IS)	17:01	1021	Auto	2,297,973	970,904
3	chrysene d-12 (IS)	22:46	1366	Auto	1,637,779	591,336
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,634,114	1,962,586
5	acenaphthene d-10 (SURR)	12:22	742	Auto	1,516,015	890,887
6	phenanthrene d-10 (SURR)	17:01	1021	Auto	2,297,973	970,904
7	chrysene d-12 (SURR)	22:46	1366	Auto	1,637,779	591,336
8	1,3-dimethyl-2-nitrobenzene	8:02	482	Auto	946,228	523,438
9	triphenyl phosphate (S)	22:10	1330	Auto	1,245,491	418,325
10	perylene d-12 (SURR)	27:09	1629	Auto	714,761	123,949
11	hexachlorocyclopentadiene	10:13	613	Auto	1,099	1,099
12	bis(2-ethylhexyl)adipate	22:11	1331	Auto	27,607	2,352
13	bis(2-ethylhexyl)phthalate	23:04	1384	Auto	376,418	102,296

reversed IS/SURR's

Quantitation Report

Quanfile: 24949B

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 24949

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	742	12:22	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1021	17:01	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1366	22:46	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	742	12:22	BV	2.798	UG/L
6	phenanthrene d-10(SURR	A	1021	17:01	BV	2.821	UG/L
7	chrysene d-12 (SURR)	A	1366	22:46	BB	2.792	UG/L
8	1,3-dimethyl-2-nitrobe	A	482	8:02	BV	5.028	UG/L
9	triphenyl phosphate (S	A	1330	22:10	BB	5.774	UG/L
10	perylene d-12 (SURR)	A	1629	27:09	BB	3.216	UG/L
11	hexachlorocyclopentadi	A	613	10:13	BB	0.016	UG/L
16	bis(2-ethylhexyl)adipa	A	1331	22:11	MM	0.157	UG/L
17	bis(2-ethylhexyl)phtha	A	1384	23:04	VB	0.689	UG/L

Comments for Sample AD24949 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-28-2001 Time: 14:24:43

Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.