


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/24/2001 Time: 15:00:44

Sample: AD25038
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/24/01 14:59 Ended: 11/8/01 14:59 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.92		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.6		1.0
21	Surr_Perylene-d12		3.3		1.0

Chromatogram Plot File: C:\25038H Date: Nov-09-1991 01:31:47
Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 25038
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 4934582



Integration Report Quanfile: 25038H Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 25038
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:22	742	Auto	1,582,352	1,026,804
2	phenanthrene d-10 (IS)	17:01	1021	Auto	2,493,182	956,931
3	chrysene d-12 (IS)	22:46	1366	Auto	1,666,225	598,991
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,310,410	1,556,095
5	acenaphthene d-10 (SURR)	12:22	742	Auto	1,582,352	1,026,804
6	phenanthrene d-10 (SURR)	17:01	1021	Auto	2,493,182	956,931
7	chrysene d-12 (SURR)	22:46	1366	Auto	1,666,225	598,991
8	1,3-dimethyl-2-nitrobenzene	8:02	482	Auto	1,008,814	543,750
9	triphenyl phosphate (S)	22:11	1331	Auto	1,217,881	422,732
10	perylene d-12 (SURR)	27:09	1629	Auto	746,208	125,075
11	hexachlorocyclopentadiene	10:13	613	Auto	1,156	1,156
12	bis(2-ethylhexyl)adipate	22:10	1330	Auto	25,822	2,987
13	bis(2-ethylhexyl)phthalate	23:04	1384	Auto	511,301	142,089
14	butachlor	20:29	1229	Auto	0	0
15	4,4'-dichlorodiphenyl ether	20:50	1250	Auto	1,040	1,040

reversed IS/ surr's

Quantitation Report Quanfile: 25038H Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/08/01; SAMPLE 25038
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	742	12:22	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1021	17:01	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1366	22:46	BV	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	742	12:22	BV	3.206	UG/L
6	phenanthrene d-10 (SURR)	A	1021	17:01	BV	3.359	UG/L
7	chrysene d-12 (SURR)	A	1366	22:46	BV	3.119	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	482	8:02	BB	5.136	UG/L
9	triphenyl phosphate (S)	A	1331	22:11	BB	5.550	UG/L
10	perylene d-12 (SURR)	A	1629	27:09	BB	3.301	UG/L
11	hexachlorocyclopentadiene	A	613	10:13	BB	0.016	UG/L
16	bis(2-ethylhexyl)adipate	A	1330	22:10	MM	0.144	UG/L
17	bis(2-ethylhexyl)phthalate	A	1384	23:04	VB	0.921	UG/L
27	butachlor	A	1229	20:29	BB	0.001	UG/L
28	4,4'-dichlorodiphenyl ether	A	1250	20:50	BB	0.004	UG/L

4 . . *

Comments for Sample AD25038 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-28-2001 Time: 14:30:12

Recoveries for 5 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.