


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/29/2001 Time: 09:28:22

Sample: AD25039
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/25/01 09:02 Ended: 11/9/01 09:02 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.3		1.0
20	Surr_Triphenyl phosphate		5.3		1.0
21	Surr_Perylene-d12		3.2		1.0

Chromatogram Plot File: C:\25039A Date: Nov-09-1991 20:34:20
Comment: BOHLK; METHOD 525; INST 204; 11/09/01; SAMPLE 25039
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 5908513



Integration Report

Quanfile: 25039A

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/09/01; SAMPLE 25039

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:21	741	Auto	1,903,145	1,217,771
2	phenanthrene d-10 (IS)	17:00	1020	Auto	3,044,071	1,338,385
3	chrysene d-12 (IS)	22:45	1365	Auto	2,126,781	838,162
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,545,297	1,752,857
5	acenaphthene d-10 (SURR)	12:21	741	Auto	1,903,145	1,217,771
6	phenanthrene d-10 (SURR)	17:00	1020	Auto	3,044,071	1,338,385
7	chrysene d-12 (SURR)	22:45	1365	Auto	2,126,781	838,162
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	1,242,029	694,505
9	triphenyl phosphate (S)	22:10	1330	Auto	1,494,302	505,796
10	perylene d-12 (SURR)	27:07	1627	Auto	921,862	157,648
11	hexachlorocyclopentadiene	10:12	612	Auto	1,377	1,377
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	38,399	4,047
13	bis(2-ethylhexyl)phthalate	23:03	1383	Auto	403,780	102,046
14	2,6-dinitrotoluene	12:00	720	Auto	105,278	45,992
15	mollinate	13:06	786	Auto	1,793	911

reviewed IS/SURR's

Quantitation Report

Quanfile: 25039A

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/09/01; SAMPLE 25039

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	741	12:21	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1020	17:00	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1365	22:45	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1254	20:54	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	741	12:21	BB	3.600	UG/L
6	phenanthrene d-10(SURR	A	1020	17:00	BV	3.830	UG/L
7	chrysene d-12 (SURR)	A	1365	22:45	BB	3.717	UG/L
8	1,3-dimethyl-2-nitrobe	A	481	8:01	BB	5.258	UG/L
9	triphenyl phosphate (S	A	1330	22:10	BB	5.335	UG/L
10	perylene d-12 (SURR)	A	1627	27:07	BB	3.195	UG/L
11	hexachlorocyclopentadi	A	612	10:12	BB	0.016	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.168	UG/L
17	bis(2-ethylhexyl)phtha	A	1383	23:03	VB	0.569	UG/L
20	2,6-dinitrotoluene	A	720	12:00	BB	0.846	UG/L
22	molinate	A	786	13:06	MM	0.006	UG/L

Comments for Sample AD25039 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-06-2001 Time: 10:42:20

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were below their acceptance ranges.