


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 11/29/2001 Time: 10:14:57

Sample: AD25051
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 10/25/01 10:11 Ended: 11/9/01 10:11 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.85		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate	USPEC	6.7		1.0
21	Surr_Perylene-d12	LSPEC	2.6		1.0

Chromatogram Plot
 File: E:\25051L Date: Nov-10-1991 08:02:10
 Comment: BOHLK; METHOD 525; INST 204; 11/09/01; SAMPLE 25051
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 8157459



Integration Report Quanfile: 25051L Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/09/01; SAMPLE 25051
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:21	741	Auto	1,802,326	997,222
2	phenanthrene d-10 (IS)	17:01	1021	Auto	2,770,260	1,097,142
3	chrysene d-12 (IS)	22:46	1366	Auto	1,760,015	718,966
4	p-terphenyl d-14 (IS)	20:54	1254	Auto	3,395,455	2,385,697
5	acenaphthene d-10 (SURR)	12:21	741	Auto	1,802,326	997,222
6	phenanthrene d-10 (SURR)	17:01	1021	Auto	2,770,260	1,097,142
7	chrysene d-12 (SURR)	22:46	1366	Auto	1,760,015	718,966
8	1,3-dimethyl-2-nitrobenzene	8:01	481	Auto	1,130,702	510,021
9	triphenyl phosphate (S)	22:10	1330	Auto	1,550,558	709,120
10	perylene d-12 (SURR)	27:08	1628	Auto	632,206	106,394
11	hexachlorocyclopentadiene	10:12	612	Auto	975	975
12	bis(2-ethylhexyl)adipate	22:10	1330	Auto	55,732	5,585
13	bis(2-ethylhexyl)phthalate	23:04	1384	Auto	498,100	126,968
14	2,6-dinitrotoluene	12:01	721	Auto	63,049	36,725
15	4,4'-dichlorodiphenyl ether	20:50	1250	Auto	2,259	1,282

reviewed IS/ surr's

Quantitation Report

Quanfile: 25051L

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/09/01; SAMPLE 25051

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	741	12:21	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1021	17:01	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1366	22:46	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1254	20:54	BV	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	741	12:21	BB	3.560	UG/L
6	phenanthrene d-10 (SURR)	A	1021	17:01	BV	3.639	UG/L
7	chrysene d-12 (SURR)	A	1366	22:46	BB	3.212	UG/L
8	1,3-dimethyl-2-nitrobe	A	481	8:01	BV	5.054	UG/L
9	triphenyl phosphate (S	A	1330	22:10	BV	6.689	UG/L
10	perylene d-12 (SURR)	A	1628	27:08	BB	2.647	UG/L
11	hexachlorocyclopentadi	A	612	10:12	BB	0.012	UG/L
16	bis(2-ethylhexyl)adipa	A	1330	22:10	MM	0.294	UG/L
17	bis(2-ethylhexyl)phta	A	1384	23:04	VV	0.849	UG/L
20	2,6-dinitrotoluene	A	721	12:01	BB	0.538	UG/L
28	4,4'-dde	A	1250	20:50	BB	0.008	UG/L

Comments for Sample AD25051 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-06-2001 Time: 11:14:21

Recoveries for 2 of the 3 Surrogate Standards in this sample were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were below their acceptance ranges.