


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/29/2001 Time: 15:04:31

Sample: AD25262
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/26/01 15:02 Ended: 11/10/01 15:02 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.77		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.0		1.0
20	Surr_Triphenyl phosphate		5.3		1.0
21	Surr_Perylene-d12		4.0		1.0

Chromatogram Plot File: E:\25262E Date: Nov-10-1991 21:45:39
Comment: BOHLK; METHOD 525; INST 204; 11/10/01; SAMPLE 25262
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 6842661



Integration Report Quanfile: 25262E Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/10/01; SAMPLE 25262
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:22	742	Auto	1,682,822	1,010,156
2	phenanthrene d-10 (IS)	17:02	1022	Auto	2,673,685	1,063,533
3	chrysene d-12 (IS)	22:47	1367	Auto	2,415,752	1,053,424
4	p-terphenyl d-14 (IS)	20:55	1255	Auto	3,116,482	1,877,681
5	acenaphthene d-10 (SURR)	12:22	742	Auto	1,682,822	1,010,156
6	phenanthrene d-10 (SURR)	17:02	1022	Auto	2,673,685	1,063,533
7	chrysene d-12 (SURR)	22:47	1367	Auto	2,415,752	1,053,424
8	1,3-dimethyl-2-nitrobenzene	8:02	482	Auto	1,052,762	502,727
9	triphenyl phosphate (S)	22:11	1331	Auto	1,688,690	727,314
10	perylene d-12 (SURR)	27:10	1630	Auto	1,295,941	238,297
11	bis(2-ethylhexyl)adipate	22:08	1328	Auto	35,976	4,063
12	bis(2-ethylhexyl)phthalate	23:04	1384	Auto	618,821	165,040
13	2,6-dinitrotoluene	12:02	722	Auto	85,360	44,901
14	4,4'-dichlorodiphenyl ether	20:50	1250	Auto	5,075	1,914

Reviewed 12/5/01

Quantitation Report

Quanfile: 25262E

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/10/01; SAMPLE 25262

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	742	12:22	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1022	17:02	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1367	22:47	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1255	20:55	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	742	12:22	BV	3.622	UG/L
6	phenanthrene d-10(SURR)	A	1022	17:02	BV	3.827	UG/L
7	chrysene d-12 (SURR)	A	1367	22:47	BB	4.802	UG/L
8	1,3-dimethyl-2-nitrobe	A	482	8:02	BV	5.040	UG/L
9	triphenyl phosphate (S	A	1331	22:11	BB	5.307	UG/L
10	perylene d-12 (SURR)	A	1630	27:10	BB	3.954	UG/L
16	bis(2-ethylhexyl)adipa	A	1328	22:08	MM	0.139	UG/L
17	bis(2-ethylhexyl)phtha	A	1384	23:04	VV	0.768	UG/L
20	2,6-dinitrotoluene	A	722	12:02	BB	0.777	UG/L
28	4,4'-dde	A	1250	20:50	BV	0.018	UG/L

Comments for Sample AD25262 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-04-2001 Time: 13:28:52

Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard and the Spiked Sample were outside of their acceptance ranges.