


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/29/2001 Time: 16:06:17

Sample: AD25460
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/29/01 16:04 Ended: 11/13/01 16:04 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.62		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.9		1.0
20	Surr_Triphenyl phosphate		5.6		1.0
21	Surr_Perylene-d12		3.8		1.0

Chromatogram Plot File: E:\25460C Date: Nov-13-1991 20:34:06
Comment: BOHLK; METHOD 525; INST 204; 11/13/01; SAMPLE 25460
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 9377019




Integration Report Quanfile: 25460C Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/13/01; SAMPLE 25460
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:17	737	Auto	2,144,147	1,406,884
2	phenanthrene d-10 (IS)	16:55	1015	Auto	3,379,356	1,256,961
3	chrysene d-12 (IS)	22:40	1360	Auto	2,214,833	920,134
4	p-terphenyl d-14 (IS)	20:50	1250	Auto	4,003,554	2,851,876
5	acenaphthene d-10 (SURR)	12:17	737	Auto	2,144,147	1,406,884
6	phenanthrene d-10 (SURR)	16:55	1015	Auto	3,379,356	1,256,961
7	chrysene d-12 (SURR)	22:40	1360	Auto	2,214,833	920,134
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	1,306,973	724,052
9	triphenyl phosphate (S)	22:05	1325	Auto	1,783,869	773,130
10	perylene d-12 (SURR)	26:57	1617	Auto	1,007,786	195,449
11	hexachlorocyclopentadiene	10:08	608	Auto	2,059	1,192
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	72,794	5,816
13	bis(2-ethylhexyl)phthalate	22:58	1378	Auto	533,647	162,112
14	2,6-dinitrotoluene	11:57	717	Auto	72,162	41,296
15	2,4-dinitrotoluene	13:04	784	Auto	2,831	1,686
16	4,4'-dichlorodiphenyl ether	20:46	1246	Auto	4,060	1,535

Reviewed 12/5/01

Quantitation Report Quanfile: 25460C Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/13/01; SAMPLE 25460
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R-Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	737	12:17	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1015	16:55	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1360	22:40	BV	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1250	20:50	BV	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	737	12:17	BB	3.524	UG/L
6	phenanthrene d-10 (SURR)	A	1015	16:55	BV	3.622	UG/L
7	chrysene d-12 (SURR)	A	1360	22:40	BV	3.559	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	478	7:58	BV	4.894	UG/L
9	triphenyl phosphate (S)	A	1325	22:05	BB	5.618	UG/L
10	perylene d-12 (SURR)	A	1617	26:57	BB	3.770	UG/L
11	hexachlorocyclopentadiene	A	608	10:08	BB	0.021	UG/L
16	bis(2-ethylhexyl)adipate	A	1322	22:02	MM	0.273	UG/L
17	bis(2-ethylhexyl)phthalate	A	1378	22:58	BV	0.620	UG/L
20	2,6-dinitrotoluene	A	717	11:57	BB	0.520	UG/L
21	2,4-dinitrotoluene	A	784	13:04	BB	0.019	UG/L
28	4,4'-dichlorodiphenyl ether	A	1246	20:46	BB	0.014	UG/L



Comments for Sample AD25460 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-07-2001 Time: 11:45:49

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges.